

# PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/19/2024

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:35  
In Date: 12/18/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:27  
Out Date: 12/19/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB133993

| Lab ID   | Client SampleID             | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|-----------------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| P5316-01 | TT-304-IDWSO-20241217-1     | 1      | 1.15           | 8.64         | 9.79                    | 7.08                      | 68.6    |                 |
| P5316-02 | TT-304-IDWSO-20241217-2     | 2      | 1.19           | 8.62         | 9.81                    | 7.6                       | 74.4    |                 |
| P5316-03 | TT-304-IDWSO-20241217-3     | 3      | 1.16           | 8.82         | 9.98                    | 8.11                      | 78.8    |                 |
| P5316-04 | TT-304-IDWSO-20241217-4     | 4      | 1.16           | 8.58         | 9.74                    | 7.36                      | 72.3    |                 |
| P5321-01 | SS055P-SD06-121324-00       | 5      | 1.16           | 8.61         | 9.77                    | 7.14                      | 69.5    |                 |
| P5321-02 | SS050P-SD21-121324-00       | 6      | 1.13           | 8.85         | 9.98                    | 8.23                      | 80.2    |                 |
| P5321-03 | SS050P-SD22-121724-00       | 7      | 1.18           | 8.77         | 9.95                    | 8.25                      | 80.6    |                 |
| P5321-04 | SS050P-SD20-121724-00       | 8      | 1.14           | 8.82         | 9.96                    | 6.56                      | 61.5    |                 |
| P5330-01 | TP-5                        | 9      | 1.17           | 8.54         | 9.71                    | 9.33                      | 95.6    |                 |
| P5330-02 | TP-5-EPH                    | 10     | 1.11           | 8.78         | 9.89                    | 9.5                       | 95.6    |                 |
| P5330-03 | TP-5-VOC                    | 11     | 1.19           | 8.61         | 9.8                     | 9.44                      | 95.8    |                 |
| P5337-01 | ETGI-357                    | 24     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | CONCRETE sample |
| P5338-01 | USED-OIL-DEBRIS             | 25     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | oily-debris     |
| P5338-02 | USED-MOBILE-TRANSFORMER-OIL | 26     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | oil sample      |
| P5339-01 | TR-06-12182024              | 27     | 1.16           | 8.77         | 9.93                    | 8.74                      | 86.4    |                 |
| P5339-02 | TR-06-12182024-E2           | 28     | 1.17           | 8.82         | 9.99                    | 9.19                      | 90.9    |                 |
| P5341-02 | STORMWATER-SOLID-COMP       | 12     | 1.15           | 8.81         | 9.96                    | 5.91                      | 54.0    | sludge sample   |
| P5342-01 | CHRT26634                   | 13     | 1.14           | 8.40         | 9.54                    | 7.05                      | 70.4    |                 |
| P5342-02 | CHRT26634-E2                | 14     | 1.16           | 8.75         | 9.91                    | 7.00                      | 66.7    |                 |
| P5342-03 | HT2651                      | 15     | 1.16           | 8.75         | 9.91                    | 8.91                      | 88.6    |                 |
| P5342-04 | HT2651-E2                   | 16     | 1.19           | 8.50         | 9.69                    | 8.82                      | 89.8    |                 |
| P5342-05 | RB21198                     | 17     | 1.17           | 8.80         | 9.97                    | 8.09                      | 78.6    |                 |
| P5342-06 | RB21198-E2                  | 18     | 1.12           | 8.87         | 9.99                    | 7.66                      | 73.7    |                 |
| P5343-01 | VNJ210                      | 19     | 1.14           | 8.63         | 9.77                    | 8.08                      | 80.4    |                 |
| P5343-02 | VNJ210-E2                   | 20     | 1.15           | 8.81         | 9.96                    | 8.9                       | 88.0    |                 |
| P5343-03 | VNJ281                      | 21     | 1.17           | 8.53         | 9.7                     | 8.55                      | 86.5    |                 |



PERCENT SOLID

Supervisor: Iwona  
Analyst: jignesh  
Date: 12/19/2024

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:35  
In Date: 12/18/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:27  
Out Date: 12/19/2024  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB133993

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| P5343-04 | VNJ281-E2       | 22     | 1.13            | 8.72          | 9.85                     | 8.83                       | 88.3    |          |
| P5343-05 | OILY-RAGS-274   | 23     | 1.14            | 8.58          | 9.72                     | 9.22                       | 94.2    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

133993

WorkList Name : %1-121824

WorkList ID : 186438

Department : Wet-Chemistry

Date : 12-18-2024 12:12:13

| Sample   | Customer Sample         | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|-------------------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| P5316-01 | TT-304-IDWSO-20241217-1 | Solid  | Percent Solids | Cool 4 deg C | TETR06   | L51                         | 12/17/2024   | Chemtech -SO |
| P5316-02 | TT-304-IDWSO-20241217-2 | Solid  | Percent Solids | Cool 4 deg C | TETR06   | L51                         | 12/17/2024   | Chemtech -SO |
| P5316-03 | TT-304-IDWSO-20241217-3 | Solid  | Percent Solids | Cool 4 deg C | TETR06   | L51                         | 12/17/2024   | Chemtech -SO |
| P5316-04 | TT-304-IDWSO-20241217-4 | Solid  | Percent Solids | Cool 4 deg C | TETR06   | L51                         | 12/17/2024   | Chemtech -SO |
| P5321-01 | SS055P-SD06-121324-00   | Solid  | Percent Solids | Cool 4 deg C | WEST04   | L41                         | 12/17/2024   | Chemtech -SO |
| P5321-02 | SS050P-SD21-121324-00   | Solid  | Percent Solids | Cool 4 deg C | WEST04   | L41                         | 12/13/2024   | Chemtech -SO |
| P5321-03 | SS050P-SD22-121724-00   | Solid  | Percent Solids | Cool 4 deg C | WEST04   | L41                         | 12/13/2024   | Chemtech -SO |
| P5321-04 | SS050P-SD20-121724-00   | Solid  | Percent Solids | Cool 4 deg C | WEST04   | L41                         | 12/17/2024   | Chemtech -SO |
| P5330-01 | TP-5                    | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | L51                         | 12/17/2024   | Chemtech -SO |
| P5330-02 | TP-5-EPH                | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | L51                         | 12/18/2024   | Chemtech -SO |
| P5330-03 | TP-5-VOC                | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | L51                         | 12/18/2024   | Chemtech -SO |
| P5337-01 | ETGI-357                | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 12/18/2024   | Chemtech -SO |
| P5338-01 | USED-OIL-DEBRIS         | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 12/18/2024   | Chemtech -SO |
| P5338-02 | USED-MOBILE-TRANSFORME  | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N11                         | 12/18/2024   | Chemtech -SO |
| P5339-01 | TR-06-12182024          | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | N12                         | 12/18/2024   | Chemtech -SO |
| P5339-02 | TR-06-12182024-E2       | Solid  | Percent Solids | Cool 4 deg C | PSEG05   | N12                         | 12/18/2024   | Chemtech -SO |
| P5341-02 | STORMWATER-SOLID-COMP   | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5342-01 | CHRT26634               | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5342-02 | CHRT26634-E2            | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5342-03 | HT2651                  | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5342-04 | HT2651-E2               | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |

Date/Time 12-18-24 16:30

Raw Sample Received by: YB cody

Raw Sample Relinquished by: CSM

Date/Time 12-18-24 17:40

Raw Sample Received by: CSM

Raw Sample Relinquished by: YB cody

# WORKLIST(Hardcopy Internal Chain)

133993

WorkList Name : %1-121824      WorkList ID : 186438      Department : Wet-Chemistry      Date : 12-18-2024 12:12:13

| Sample   | Customer Sample | Matrix | Test           | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method       |
|----------|-----------------|--------|----------------|--------------|----------|-----------------------------|--------------|--------------|
| P5342-05 | RB21198         | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5342-06 | RB21198-E2      | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5343-01 | VNJ210          | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5343-02 | VNJ210-E2       | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5343-03 | VNJ281          | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5343-04 | VNJ281-E2       | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |
| P5343-05 | OILY-RAGS-274   | Solid  | Percent Solids | Cool 4 deg C | PSEG03   | N12                         | 12/18/2024   | Chemtech -SO |

Date/Time 12-18-24 16:30  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]

Date/Time 12-18-24 17:40  
 Raw Sample Received by: [Signature]  
 Raw Sample Relinquished by: [Signature]