

## **CASE NARRATIVE**

**AECOM Technical Services, Inc.**

**Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258**

**Project # N/A**

**Chemtech Project # P5317**

**Test Name: VOCMS Group1**

### **A. Number of Samples and Date of Receipt:**

25 Water samples were received on 12/17/2024.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group1 was based on method 8260D.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds .

The MSD recoveries met the acceptable requirements .

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Initial Calibration method (82X121124W.M) for t-1,3-Dichloropropene, Dibromochloromethane, Bromoform these compounds are passing on Quadratic Regression.

The Continuous Calibration File ID VX044397.D met the requirements except for Chloroethane is failing high but no positive hit in associate sample therefore no corrective action taken.

The Continuous Calibration File ID VX044438.D met the requirements except for Methylcyclohexane is failing high but no positive hit in associate sample therefore no corrective action taken.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The Continuous Calibration File ID VX044466.D met the requirements except for Tetrachloroethene failing marginally high and Methylcyclohexane is failing high but no positive hit in associate sample therefore no corrective action taken.  
The Tuning criteria met requirements.

Samples RE132D7-20241217 was diluted due to samples past history of high concentration of Trichloroethene.

Samples RW4-20241216, RE132D5-20241217 were diluted due to high concentrations.

**E. Additional Comments:**

The not QT review data is reported in the Miscellaneous.  
Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_