

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG No.: P5317

Instrument ID: BNA_N Calibration Date/Time: 12/20/2024 19:33

Lab File ID: BN035749.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.622		-0.6	20.0
Fluoranthene-d10	1.134	1.074		-5.3	20.0
2-Fluorophenol	1.001	1.021		2.0	20.0
Phenol-d6	1.204	1.178		-2.2	20.0
Nitrobenzene-d5	0.244	0.294		20.5	20.0
2-Fluorobiphenyl	1.512	1.560		3.2	20.0
2,4,6-Tribromophenol	0.284	0.235		-17.3	20.0
Terphenyl-d14	0.789	0.891		12.9	20.0
1,4-Dioxane	0.382	0.445		16.5	20.0

All other compounds must meet a minimum RRF of 0.010.