

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG No.: P5317

Instrument ID: BNA_N Calibration Date/Time: 12/26/2024 19:41

Lab File ID: BN035861.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.551		-12.0	50.0
Fluoranthene-d10	1.134	0.922		-18.7	50.0
2-Fluorophenol	1.001	0.931		-7.0	50.0
Phenol-d6	1.204	1.022		-15.1	50.0
Nitrobenzene-d5	0.244	0.306		25.4	50.0
2-Fluorobiphenyl	1.512	1.480		-2.1	50.0
2,4,6-Tribromophenol	0.284	0.225		-20.8	50.0
Terphenyl-d14	0.789	0.898		13.8	50.0
1,4-Dioxane	0.382	0.394		3.1	50.0

All other compounds must meet a minimum RRF of 0.010.