

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	12/17/24
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	12/17/24
Client Sample ID:	TT158S1-20241217		SDG No.:	P5317
Lab Sample ID:	P5317-09		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035755.D	1	12/18/24 11:40	12/20/24 23:09	PB165723

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	1.20		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.43		30 - 150		108%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.46		30 - 150		115%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.45	*	55 - 111		112%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		97%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.77	*	58 - 132		192%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2810	7.272				
1146-65-2	Naphthalene-d8	7440	10.02				
15067-26-2	Acenaphthene-d10	5100	13.924				
1517-22-2	Phenanthrene-d10	10400	16.698				
1719-03-5	Chrysene-d12	8560	20.956				
1520-96-3	Perylene-d12	8470	23.041				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products