

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG NO.: P5317

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/23/2024

Lab File ID: BN035794.D Time Analyzed: 22:16

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	8845	16.698	7163	20.947	7277	23.032
UPPER LIMIT	17690	17.198	14326	21.447	14554	23.532
LOWER LIMIT	4422.5	16.198	3581.5	20.447	3638.5	22.532
EPA SAMPLE NO.						
01 DUP-04-20241216	11883	16.70	10654	20.95	10068	23.03
02 RW8-MW01S-20241216	13726	16.70	11081	20.95	9977	23.03

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.