

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: P5317

SDG NO.: P5317

Lab File ID: BN035844.D

DFTPP Injection Date: 12/26/2024

Instrument ID: BNA\_N

DFTPP Injection Time: 09:19

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	24
68	Less than 2.0% of mass 69	0.2 ( 0.6 ) 1
69	Mass 69 relative abundance	32.9
70	Less than 2.0% of mass 69	0.2 ( 0.5 ) 1
127	10.0 - 80.0% of mass 198	40
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	27.1
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	11.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	14.1 (20.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035845.D	12/26/2024	09:58
PB165724BL	PB165724BL	BN035846.D	12/26/2024	10:35
PB165723BS	PB165723BS	BN035851.D	12/26/2024	13:43
PB165724BSD	PB165724BSD	BN035853.D	12/26/2024	14:54
RE132D6-20241217DL	P5317-21DL	BN035855.D	12/26/2024	16:06
SSTDCCC0.4EC	SSTDCCC0.4	BN035861.D	12/26/2024	19:41