

Surrogate Summary

SDG No.: **P5317**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5317-01	RE114D2-20241216	1,2-Dichloroethane-d4	50	44.7	89	81	118
		Dibromofluoromethane	50	47.9	96	80	119
		Toluene-d8	50	53.1	106	89	112
		4-Bromofluorobenzene	50	48.7	97	85	114
P5317-02	RW4-20241216	1,2-Dichloroethane-d4	50	46.4	93	81	118
		Dibromofluoromethane	50	47.6	95	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	46.5	93	85	114
P5317-02DL	RW4-20241216DL	1,2-Dichloroethane-d4	50	42.7	85	81	118
		Dibromofluoromethane	50	47.4	95	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	47.8	96	85	114
P5317-03	TT191D1-20241216	1,2-Dichloroethane-d4	50	42.4	85	81	118
		Dibromofluoromethane	50	47.2	94	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	45.9	92	85	114
P5317-04	TT191D2-20241216	1,2-Dichloroethane-d4	50	44.5	89	81	118
		Dibromofluoromethane	50	47.4	95	80	119
		Toluene-d8	50	51.5	103	89	112
		4-Bromofluorobenzene	50	48.0	96	85	114
P5317-05	TT163S1-20241216	1,2-Dichloroethane-d4	50	45.0	90	81	118
		Dibromofluoromethane	50	48.1	96	80	119
		Toluene-d8	50	52.9	106	89	112
		4-Bromofluorobenzene	50	46.8	94	85	114
P5317-06	DUP-04-20241216	1,2-Dichloroethane-d4	50	42.8	86	81	118
		Dibromofluoromethane	50	48.7	97	80	119
		Toluene-d8	50	52.6	105	89	112
		4-Bromofluorobenzene	50	48.4	97	85	114
P5317-07	RW8-MW01S-20241216	1,2-Dichloroethane-d4	50	45.5	91	81	118
		Dibromofluoromethane	50	46.9	94	80	119
		Toluene-d8	50	52.0	104	89	112
		4-Bromofluorobenzene	50	46.3	93	85	114
P5317-08	RW8-MW01D1-20241216	1,2-Dichloroethane-d4	50	45.0	90	81	118
		Dibromofluoromethane	50	46.3	93	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	45.9	92	85	114
P5317-09	TT158S1-20241217	1,2-Dichloroethane-d4	50	45.4	91	81	118
		Dibromofluoromethane	50	47.3	95	80	119
		Toluene-d8	50	51.6	103	89	112
		4-Bromofluorobenzene	50	44.6	89	85	114
P5317-10	TT158I1-20241217	1,2-Dichloroethane-d4	50	44.2	88	81	118
		Dibromofluoromethane	50	47.8	95	80	119
		Toluene-d8	50	51.1	102	89	112
		4-Bromofluorobenzene	50	45.6	91	85	114
P5317-11	DUP-05-20241217	1,2-Dichloroethane-d4	50	44.9	90	81	118
		Dibromofluoromethane	50	49.6	99	80	119
		Toluene-d8	50	55.6	111	89	112
		4-Bromofluorobenzene	50	53.1	106	85	114
P5317-12	TT172S1-20241217	1,2-Dichloroethane-d4	50	45.5	91	81	118
		Dibromofluoromethane	50	47.9	96	80	119

Surrogate Summary

SDG No.: **P5317**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5317-12	TT172S1-20241217	Toluene-d8	50	52.6	105	89	112
		4-Bromofluorobenzene	50	47.8	96	85	114
P5317-13	RW9-MW01S-20241217	1,2-Dichloroethane-d4	50	46.3	93	81	118
		Dibromofluoromethane	50	46.9	94	80	119
		Toluene-d8	50	51.4	103	89	112
		4-Bromofluorobenzene	50	47.9	96	85	114
P5317-14	RW9-MW01D1-20241217	1,2-Dichloroethane-d4	50	46.1	92	81	118
		Dibromofluoromethane	50	47.3	95	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	46.3	93	85	114
P5317-15	RW9-MW01D2-20241217	1,2-Dichloroethane-d4	50	45.7	91	81	118
		Dibromofluoromethane	50	48.0	96	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	45.0	90	85	114
P5317-16	RW9-MW01D3-20241217	1,2-Dichloroethane-d4	50	45.4	91	81	118
		Dibromofluoromethane	50	47.5	95	80	119
		Toluene-d8	50	51.6	103	89	112
		4-Bromofluorobenzene	50	44.8	90	85	114
P5317-17	DUP-06-20241217	1,2-Dichloroethane-d4	50	44.3	89	81	118
		Dibromofluoromethane	50	50.2	100	80	119
		Toluene-d8	50	54.4	109	89	112
		4-Bromofluorobenzene	50	50.2	100	85	114
P5317-18	RE132D5-20241217	1,2-Dichloroethane-d4	50	45.4	91	81	118
		Dibromofluoromethane	50	48.4	97	80	119
		Toluene-d8	50	52.1	104	89	112
		4-Bromofluorobenzene	50	48.4	97	85	114
P5317-19MS	RE132D5-20241217MS	1,2-Dichloroethane-d4	50	50.1	100	81	118
		Dibromofluoromethane	50	54.5	109	80	119
		Toluene-d8	50	55.2	110	89	112
		4-Bromofluorobenzene	50	54.8	110	85	114
P5317-20MSD	RE132D5-20241217MSD	1,2-Dichloroethane-d4	50	49.1	98	81	118
		Dibromofluoromethane	50	52.6	105	80	119
		Toluene-d8	50	52.5	105	89	112
		4-Bromofluorobenzene	50	53.4	107	85	114
P5317-21	RE132D6-20241217	1,2-Dichloroethane-d4	50	41.4	83	81	118
		Dibromofluoromethane	50	47.8	96	80	119
		Toluene-d8	50	53.3	107	89	112
		4-Bromofluorobenzene	50	50.2	100	85	114
P5317-22	RE132D7-20241217	1,2-Dichloroethane-d4	50	42.0	84	81	118
		Dibromofluoromethane	50	50.2	100	80	119
		Toluene-d8	50	53.0	106	89	112
		4-Bromofluorobenzene	50	51.7	103	85	114
P5317-23	TB	1,2-Dichloroethane-d4	50	44.5	89	81	118
		Dibromofluoromethane	50	46.5	93	80	119
		Toluene-d8	50	51.5	103	89	112
		4-Bromofluorobenzene	50	46.2	92	85	114
P5317-24	RW8-MW01D2-20241217	1,2-Dichloroethane-d4	50	43.5	87	81	118
		Dibromofluoromethane	50	49.9	100	80	119
		Toluene-d8	50	54.4	109	89	112
		4-Bromofluorobenzene	50	51.5	103	85	114

Surrogate Summary

SDG No.: P5317

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5317-25	RW8-MW01D3-20241217	1,2-Dichloroethane-d4	50	43.5	87	81	118
		Dibromofluoromethane	50	49.2	98	80	119
		Toluene-d8	50	54.8	110	89	112
		4-Bromofluorobenzene	50	50.8	102	85	114
VX1218WBL02	VX1218WBL02	1,2-Dichloroethane-d4	50	43.6	87	81	118
		Dibromofluoromethane	50	48.3	97	80	119
		Toluene-d8	50	51.9	104	89	112
		4-Bromofluorobenzene	50	48.2	96	85	114
VX1218WBS02	VX1218WBS02	1,2-Dichloroethane-d4	50	50.5	101	81	118
		Dibromofluoromethane	50	52.7	105	80	119
		Toluene-d8	50	50.9	102	89	112
		4-Bromofluorobenzene	50	48.5	97	85	114
VX1220WBL01	VX1220WBL01	1,2-Dichloroethane-d4	50	44.1	88	81	118
		Dibromofluoromethane	50	47.5	95	80	119
		Toluene-d8	50	52.0	104	89	112
		4-Bromofluorobenzene	50	48.4	97	85	114
VX1220WBS01	VX1220WBS01	1,2-Dichloroethane-d4	50	42.7	85	81	118
		Dibromofluoromethane	50	48.5	97	80	119
		Toluene-d8	50	48.7	97	89	112
		4-Bromofluorobenzene	50	46.4	93	85	114
VX1223WBL01	VX1223WBL01	1,2-Dichloroethane-d4	50	41.8	84	81	118
		Dibromofluoromethane	50	46.7	93	80	119
		Toluene-d8	50	51.2	102	89	112
		4-Bromofluorobenzene	50	45.8	92	85	114
VX1223WBS01	VX1223WBS01	1,2-Dichloroethane-d4	50	44.0	88	81	118
		Dibromofluoromethane	50	54.2	108	80	119
		Toluene-d8	50	56.2	112	89	112
		4-Bromofluorobenzene	50	53.5	107	85	114