

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5317

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1220WBS01	Dichlorodifluoromethane	20	18.5	ug/L	93			32	152	
	Chloromethane	20	18.7	ug/L	94			50	139	
	Vinyl chloride	20	17.9	ug/L	90			58	137	
	Bromomethane	20	18.8	ug/L	94			53	141	
	Chloroethane	20	19.8	ug/L	99			60	138	
	Trichlorofluoromethane	20	15.3	ug/L	77			65	141	
	1,1,2-Trichlorotrifluoroethane	20	18.5	ug/L	93			70	136	
	1,1-Dichloroethene	20	18.9	ug/L	95			71	131	
	Acetone	100	75.7	ug/L	76			39	160	
	Carbon disulfide	20	16.4	ug/L	82			64	133	
	Methyl tert-butyl Ether	20	17.6	ug/L	88			71	124	
	Methyl Acetate	20	17.7	ug/L	89			56	136	
	Methylene Chloride	20	19.1	ug/L	96			74	124	
	trans-1,2-Dichloroethene	20	18.7	ug/L	94			75	124	
	1,1-Dichloroethane	20	18.9	ug/L	95			77	125	
	Cyclohexane	20	17.6	ug/L	88			71	130	
	2-Butanone	100	83.3	ug/L	83			56	143	
	Carbon Tetrachloride	20	18.9	ug/L	95			72	136	
	cis-1,2-Dichloroethene	20	18.0	ug/L	90			78	123	
	Bromochloromethane	20	15.8	ug/L	79			78	123	
	Chloroform	20	17.7	ug/L	89			79	124	
	1,1,1-Trichloroethane	20	17.5	ug/L	88			74	131	
	Methylcyclohexane	20	19.3	ug/L	97			72	132	
	Benzene	20	20.4	ug/L	102			79	120	
	1,2-Dichloroethane	20	19.2	ug/L	96			73	128	
	Trichloroethene	20	19.9	ug/L	100			79	123	
	1,2-Dichloropropane	20	19.1	ug/L	96			78	122	
	Bromodichloromethane	20	19.1	ug/L	96			79	125	
	4-Methyl-2-Pentanone	100	92.4	ug/L	92			67	130	
	Toluene	20	19.5	ug/L	98			80	121	
	t-1,3-Dichloropropene	20	17.9	ug/L	90			73	127	
	cis-1,3-Dichloropropene	20	19.2	ug/L	96			75	124	
	1,1,2-Trichloroethane	20	20.8	ug/L	104			80	119	
	2-Hexanone	100	90.4	ug/L	90			57	139	
	Dibromochloromethane	20	18.8	ug/L	94			74	126	
	1,2-Dibromoethane	20	20.2	ug/L	101			77	121	
	Tetrachloroethene	20	20.8	ug/L	104			74	129	
	Chlorobenzene	20	19.9	ug/L	100			82	118	
	Ethyl Benzene	20	19.2	ug/L	96			79	121	
	m/p-Xylenes	40	39.8	ug/L	100			80	121	
	o-Xylene	20	19.8	ug/L	99			78	122	
	Styrene	20	19.8	ug/L	99			78	123	
	Bromoform	20	19.3	ug/L	97			66	130	
	Isopropylbenzene	20	19.9	ug/L	100			72	131	
	1,1,2,2-Tetrachloroethane	20	19.6	ug/L	98			71	121	
	1,3-Dichlorobenzene	20	20.1	ug/L	101			80	119	
	1,4-Dichlorobenzene	20	20.0	ug/L	100			79	118	
	1,2-Dichlorobenzene	20	19.8	ug/L	99			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5317

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX044441.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX1220WBS01	1,2-Dibromo-3-Chloropropane	20	16.5	ug/L	83			62	128	
	1,2,4-Trichlorobenzene	20	19.5	ug/L	98			69	130	
	1,2,3-Trichlorobenzene	20	19.4	ug/L	97			69	129	