

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>CHEMTECH</u>			Contract:	<u>AECO15</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5317</u>	SAS No.:	<u>P5317</u>	SDG NO.:	<u>P5317</u>
Lab File ID:	<u>VX044397.D</u>			Date Analyzed:	<u>12/18/2024</u>		
Instrument ID:	<u>MSVOA_X</u>			Time Analyzed:	<u>21:48</u>		
GC Column:	DB-624UI	ID:	0.18 (mm)	Heated Purge:	(Y/N)	N	

	IS4 AREA #	RT #				
12 HOUR STD	94679	12.018				
UPPER LIMIT	189358	12.518				
LOWER LIMIT	47339.5	11.518				
EPA SAMPLE NO.						
RE114D2-20241216	72683	12.02				
RW4-20241216	70176	12.02				
TT191D2-20241216	73178	12.02				
TT163S1-20241216	67456	12.02				
RW8-MW01S-20241216	66576	12.02				
RW8-MW01D1-20241216	71479	12.02				
TT158S1-20241217	65660	12.02				
TT158I1-20241217	64056	12.02				
TT172S1-20241217	68296	12.02				
RW9-MW01S-20241217	73713	12.02				
RW9-MW01D1-20241217	74025	12.02				
RW9-MW01D2-20241217	62374	12.02				
RW9-MW01D3-20241217	67520	12.02				
RE132D5-20241217	71901	12.02				
RE132D5-20241217MS	89259	12.02				
RE132D5-20241217MSD	88743	12.02				
TB	68420	12.02				
VX1218WBL02	77360	12.02				
VX1218WBS02	86763	12.02				

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG NO.: P5317
 Lab File ID: VX044397.D Date Analyzed: 12/18/2024
 Instrument ID: MSVOA_X Time Analyzed: 21:48
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	94679	12.018				
UPPER LIMIT	189358	12.518				
LOWER LIMIT	47339.5	11.518				
EPA SAMPLE NO.						

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.