

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG No.: P5317

Instrument ID: MSVOA_X Calibration Date/Time: 12/23/2024 19:49

Lab File ID: VX044492.D Init. Calib. Date(s): 12/11/2024 12/11/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.773	0.700		-9.44	50
Chloromethane	0.739	0.678	0.1	-8.25	50
Vinyl Chloride	0.770	0.689		-10.52	50
Bromomethane	0.546	0.468		-14.29	50
Chloroethane	0.480	0.442		-7.92	50
Trichlorofluoromethane	1.508	1.069		-29.11	50
1,1,2-Trichlorotrifluoroethane	0.610	0.551		-9.67	50
1,1-Dichloroethene	0.560	0.531		-5.18	50
Acetone	0.357	0.251		-29.69	50
Carbon Disulfide	1.065	0.983		-7.7	50
Methyl tert-butyl Ether	2.187	1.845		-15.64	50
Methyl Acetate	0.922	0.733		-20.5	50
Methylene Chloride	0.671	0.623		-7.15	50
trans-1,2-Dichloroethene	0.596	0.544		-8.73	50
1,1-Dichloroethane	1.172	1.059	0.1	-9.64	50
Cyclohexane	0.946	0.810		-14.38	50
2-Butanone	0.508	0.382		-24.8	50
Carbon Tetrachloride	0.482	0.477		-1.04	50
cis-1,2-Dichloroethene	0.756	0.679		-10.19	50
Bromochloromethane	0.575	0.456		-20.7	50
Chloroform	1.316	1.143		-13.15	50
1,1,1-Trichloroethane	1.093	0.986		-9.79	50
Methylcyclohexane	0.535	0.537		0.37	50
Benzene	1.359	1.392		2.43	50
1,2-Dichloroethane	0.560	0.534		-4.64	50
Trichloroethene	0.339	0.355		4.72	50
1,2-Dichloropropane	0.347	0.352		1.44	50
Bromodichloromethane	0.465	0.489		5.16	50
4-Methyl-2-Pentanone	0.556	0.477		-14.21	50
Toluene	0.853	0.862		1.05	50
t-1,3-Dichloropropene	0.455	0.469		3.08	50
cis-1,3-Dichloropropene	0.507	0.529		4.34	50
1,1,2-Trichloroethane	0.339	0.347		2.36	50
2-Hexanone	0.403	0.346		-14.14	50
Dibromochloromethane	0.326	0.351		7.67	50
1,2-Dibromoethane	0.345	0.357		3.48	50
Tetrachloroethene	0.332	0.362		9.04	50
Chlorobenzene	1.071	1.103	0.3	2.99	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG No.: P5317
 Instrument ID: MSVOA_X Calibration Date/Time: 12/23/2024 19:49
 Lab File ID: VX044492.D Init. Calib. Date(s): 12/11/2024 12/11/2024
 Heated Purge: (Y/N) N Init. Calib. Time(s): 10:41 13:00
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.851	1.878		1.46	50
m/p-Xylenes	0.679	0.714		5.16	50
o-Xylene	0.687	0.708		3.06	50
Styrene	1.102	1.181		7.17	50
Bromoform	0.219	0.245	0.1	11.87	50
Isopropylbenzene	3.828	3.728		-2.61	50
1,1,2,2-Tetrachloroethane	1.329	1.216	0.3	-8.5	50
1,3-Dichlorobenzene	1.621	1.672		3.15	50
1,4-Dichlorobenzene	1.667	1.677		0.6	50
1,2-Dichlorobenzene	1.684	1.667		-1.01	50
1,2-Dibromo-3-Chloropropane	0.258	0.216		-16.28	50
1,2,4-Trichlorobenzene	0.923	0.999		8.23	50
1,2,3-Trichlorobenzene	0.979	1.021		4.29	50
1,2-Dichloroethane-d4	0.877	0.643		-26.68	50
Dibromofluoromethane	0.346	0.316		-8.67	50
Toluene-d8	1.168	1.039		-11.05	50
4-Bromofluorobenzene	0.408	0.358		-12.26	50

All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.