



Lab Name:	<u>CHEMTECH</u>			Contract:	<u>AECO15</u>		
Lab Code:	<u>CHEM</u>	Case No.:	<u>P5317</u>	SAS No.:	<u>P5317</u>	SDG NO.:	<u>P5317</u>
Lab File ID:	<u>VX044397.D</u>			Date Analyzed:	<u>12/18/2024</u>		
Instrument ID:	<u>MSVOA_X</u>			Time Analyzed:	<u>21:48</u>		
GC Column:	DB-624UI	ID:	0.18 (mm)	Heated Purge:	(Y/N)	N	

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	135217	5.54	239936	6.76	204282	10.06
UPPER LIMIT	270434	6.044	479872	7.257	408564	10.555
LOWER LIMIT	67608.5	5.044	119968	6.257	102141	9.555
EPA SAMPLE NO.						
RE114D2-20241216	107762	5.55	195811	6.76	176405	10.06
RW4-20241216	108595	5.55	199832	6.76	172545	10.06
TT191D2-20241216	108502	5.55	198842	6.76	177204	10.06
TT163S1-20241216	107084	5.55	188252	6.76	168467	10.06
RW8-MW01S-20241216	108582	5.54	195784	6.76	169226	10.06
RW8-MW01D1-20241216	111969	5.55	205674	6.76	179449	10.06
TT158S1-20241217	107112	5.55	197035	6.76	171282	10.06
TT158I1-20241217	108444	5.54	193059	6.76	169424	10.06
TT172S1-20241217	106531	5.54	190393	6.76	168278	10.06
RW9-MW01S-20241217	106718	5.55	197494	6.76	178573	10.06
RW9-MW01D1-20241217	111175	5.55	207120	6.76	180812	10.06
RW9-MW01D2-20241217	104505	5.55	188770	6.76	164946	10.06
RW9-MW01D3-20241217	107842	5.55	195942	6.76	171336	10.06
RE132D5-20241217	104497	5.54	189474	6.76	170677	10.06
RE132D5-20241217MS	120579	5.55	213617	6.76	190465	10.06
RE132D5-20241217MSD	116949	5.55	212543	6.76	188325	10.06
TB	109432	5.55	198449	6.76	170843	10.06
VX1218WBL02	114674	5.55	206506	6.76	183436	10.06
VX1218WBS02	128991	5.55	233001	6.76	200137	10.06

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG NO.: P5317
 Lab File ID: VX044397.D Date Analyzed: 12/18/2024
 Instrument ID: MSVOA_X Time Analyzed: 21:48
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	135217	5.54	239936	6.76	204282	10.06
UPPER LIMIT	270434	6.044	479872	7.257	408564	10.555
LOWER LIMIT	67608.5	5.044	119968	6.257	102141	9.555
EPA SAMPLE NO.						

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.