

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG NO.: P5317
Lab File ID: VX044540.D Date Analyzed: 12/30/2024
Instrument ID: MSVOA_X Time Analyzed: 13:07
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	137969	5.54	223553	6.76	195826	10.06
UPPER LIMIT	275938	6.044	447106	7.257	391652	10.555
LOWER LIMIT	68984.5	5.044	111777	6.257	97913	9.555
EPA SAMPLE NO.						
RE132D5-20241217DL	117987	5.55	200283	6.76	179426	10.06
VX1231WBL01	117464	5.55	200538	6.76	172503	10.06
VX1231WBS01	153746	5.54	254154	6.76	219785	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.