

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5317 SAS No.: P5317 SDG NO.: P5317
 Lab File ID: VX044466.D Date Analyzed: 12/23/2024
 Instrument ID: MSVOA_X Time Analyzed: 08:49
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	59280	12.018				
UPPER LIMIT	118560	12.518				
LOWER LIMIT	29640	11.518				
EPA SAMPLE NO.						
DUP-04-20241216	92708	12.02				
DUP-05-20241217	95898	12.02				
DUP-06-20241217	84907	12.02				
RE132D6-20241217	102248	12.02				
RE132D7-20241217	100422	12.02				
RW8-MW01D2-20241217	94135	12.02				
RW8-MW01D3-20241217	89186	12.02				
VX1223WBL01	97837	12.02				
VX1223WBS03	93712	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.