

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5320

Project: Monthly 2024

Run No.: LB133994

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133994BS							
BOD5	mg/L	198	198		100	1	84.6-115.4	12/18/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5320

Project: Monthly 2024

Run No.: LB134009

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134009BS							
TSS	mg/L	550	594		108	1	90-110	12/19/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5320

Project: Monthly 2024

Run No.: LB134029

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134029BS							
TPH	mg/L	20.0	16.7		84	1	78-114	12/20/2024

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P5320

Project: Monthly 2024

Run No.: LB134029

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134029BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	12/20/2024