

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5321

Project: Joint Base McGuire Dix Lakehurst JBMDL - DoD

RunNo.: LB134031

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	934	1000	93	90-110	11/12/2024
Sample ID:	CCV1						
TOC		mg/L	968	1000	97	90-110	12/20/2024
Sample ID:	CCV2						
TOC		mg/L	1000	1000	100	90-110	12/20/2024
Sample ID:	CCV3						
TOC		mg/L	947	1000	95	90-110	12/23/2024
Sample ID:	CCV4						
TOC		mg/L	1020	1000	102	90-110	12/23/2024
Sample ID:	CCV5						
TOC		mg/L	999	1000	100	90-110	12/23/2024