

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LEO Ingwer, Inc. **SDG No.:** P5327
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** P5327 **SAS No.:** P5327
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	522	510	102	95 - 105	P	12/20/2024	16:07	LB134054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LEO Ingwer, Inc. **SDG No.:** P5327
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** P5327 **SAS No.:** P5327
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.8	20.0	109	80 - 120	P	12/20/2024	16:11	LB134054

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LEO Ingwer, Inc. **SDG No.:** P5327

Contract: LEOI01 **Lab Code:** CHEM **Case No.:** P5327 **SAS No.:** P5327

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1220	1250	97	90 - 110	P	12/20/2024	16:47	LB134054
CCV02	Copper	1260	1250	101	90 - 110	P	12/20/2024	17:29	LB134054
CCV03	Copper	1250	1250	100	90 - 110	P	12/20/2024	18:12	LB134054
CCV04	Copper	1240	1250	99	90 - 110	P	12/20/2024	19:02	LB134054
CCV05	Copper	1270	1250	102	90 - 110	P	12/20/2024	19:57	LB134054
CCV06	Copper	1220	1250	98	90 - 110	P	12/20/2024	20:47	LB134054
CCV07	Copper	1230	1250	99	90 - 110	P	12/20/2024	21:37	LB134054
CCV08	Copper	1260	1250	101	90 - 110	P	12/20/2024	23:01	LB134054
CCV09	Copper	1250	1250	100	90 - 110	P	12/20/2024	23:55	LB134054