



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: LEO Ingwer, Inc. **SDG No.:** P5327
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** P5327 **SAS No.:** P5327
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	21.8	20.0	109	40 - 160	P	12/20/2024	16:26	LB134054