

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: LEO Ingwer, Inc. **SDG No.:** P5327
Contract: LEOI01 **Lab Code:** CHEM **Case No.:** P5327 **SAS No.:** P5327
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	16.4	2.0	820	-18	22	12/20/2024	16:30	LB134054
ICSAB01	Copper	520	511	102	434	588	12/20/2024	16:34	LB134054
ICSA	Copper	18.5	2.0	925	-18	22	12/20/2024	16:38	LB134054
ICSAB	Copper	561	511	110	434	588	12/20/2024	16:43	LB134054