

Surrogate Summary

SDG No.: P5328

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5328-01	001-WILLETS-PT-BLVD(DEC)	1,2-Dichloroethane-d4	30	30.4	101	91	110
		Toluene-d8	30	27.9	93	91	112
		4-Bromofluorobenzene	30	25.1	84	63	112
P5328-02	002-35TH-AVE(DEC)	1,2-Dichloroethane-d4	30	30.1	100	91	110
		Toluene-d8	30	27.3	91	91	112
		4-Bromofluorobenzene	30	24.7	82	63	112
VN1220WBL01	VN1220WBL01	1,2-Dichloroethane-d4	30	30.5	102	91	110
		Toluene-d8	30	28.5	95	91	112
		4-Bromofluorobenzene	30	23.8	79	63	112
VN1220WBS01	VN1220WBS01	1,2-Dichloroethane-d4	30	29.4	98	91	110
		Toluene-d8	30	29.9	100	91	112
		4-Bromofluorobenzene	30	29.5	98	63	112
VN1220WBSD0	VN1220WBSD01	1,2-Dichloroethane-d4	30	30.5	102	91	110
		Toluene-d8	30	29.7	99	91	112
		4-Bromofluorobenzene	30	30.4	101	63	112