

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: P5328 SAS No.: P5328 SDG No.: P5328

Instrument ID: MSVOA_N Calibration Date/Time: 12/20/2024 10:29

Lab File ID: VN085273.D Init. Calib. Date(s): 12/20/2024 12/20/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:05 11:42

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.646	1.700		3.28	
Toluene	1.843	1.904		3.31	
Ethyl Benzene	1.939	1.922		-0.88	
m/p-Xylenes	0.728	0.752		3.3	
o-Xylene	0.693	0.695		0.29	
1,2-Dichloroethane-d4	2.639	2.656		0.64	
Toluene-d8	1.465	1.510		3.07	
4-Bromofluorobenzene	0.486	0.488		0.41	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.