

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122024\
 Data File : VN085276.D
 Acq On : 20 Dec 2024 11:42
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Dec 21 01:44:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N122024W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Dec 21 01:29:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	29871	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	167650	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	157508	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	80425	30.606	ug/l	0.00
Spiked Amount 30.000	Range 91 - 110		Recovery = 102.033%			
60) 4-Bromofluorobenzene	12.847	95	78980	30.956	ug/l	0.00
Spiked Amount 30.000	Range 63 - 112		Recovery = 103.200%			
63) Toluene-d8	10.565	98	225212	29.282	ug/l	0.00
Spiked Amount 30.000	Range 91 - 112		Recovery = 97.600%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	2.124	85	381612	152.790	ug/l	97
3) Chloromethane	2.353	50	362946	143.926	ug/l	98
4) Vinyl Chloride	2.506	62	380504	148.603	ug/l	100
5) Bromomethane	2.924	94	223224	149.127	ug/l	96
6) Chloroethane	3.100	64	232256	143.688	ug/l	91
7) Trichlorofluoromethane	3.489	101	521516	148.252	ug/l	98
8) Diethyl Ether	3.953	74	193189	159.461	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.365	101	303693	150.342	ug/l	97
10) 1,1-Dichloroethene	4.330	96	289766	157.023	ug/l	96
11) Methyl Iodide	4.583	142	378427	171.110	ug/l	99
12) Methyl Acetate	5.018	43	448482	162.865	ug/l	92
13) Acrolein	4.171	56	190706	775.174	ug/l	99
14) Acrylonitrile	5.712	53	865051	811.596	ug/l	100
15) Acetone	4.424	58	208374	774.994	ug/l	99
16) Carbon Disulfide	4.706	76	845730	150.770	ug/l	97
17) Allyl chloride	5.018	41	486116	162.845	ug/l	96
18) Methylene Chloride	5.271	84	327575	148.968	ug/l	97
19) trans-1,2-Dichloroethene	5.777	96	300282	156.451	ug/l	94
20) Diisopropyl ether	6.671	45	1090650	159.418	ug/l	95
21) 1,1-Dichloroethane	6.565	63	610620	152.673	ug/l	97
22) cis-1,2-Dichloroethene	7.483	96	357862	156.578	ug/l	98
23) tert-Butyl Alcohol	5.524	59	295956	842.524	ug/l #	100
24) Methyl tert-Butyl Ether	5.789	73	1008587	166.517	ug/l	98
25) Chloroform	7.965	83	605692	149.397	ug/l	100
26) Cyclohexane	8.253	56	502835	164.897	ug/l #	98
29) 1,1-Dichloropropene	8.371	75	433829	149.042	ug/l	100
30) 2-Butanone	7.477	43	1170394	761.746	ug/l	99
31) 2,2-Dichloropropane	7.488	77	535416	145.163	ug/l	100
32) 1,1,1-Trichloroethane	8.165	97	537681	142.095	ug/l	98
33) Carbon Tetrachloride	8.359	117	470504	142.043	ug/l	95
34) Benzene	8.606	78	1316056	143.036	ug/l	99
35) Methacrylonitrile	7.777	41	260625	153.232	ug/l	97
36) 1,2-Dichloroethane	8.665	62	463319	139.254	ug/l	100
37) Trichloroethene	9.347	130	295039	145.744	ug/l	96
38) Methylcyclohexane	9.600	83	478702	158.922	ug/l	93
39) 1,2-Dichloropropane	9.618	63	338235	144.489	ug/l	99
40) Dibromomethane	9.706	93	231145	141.625	ug/l	99
41) Bromodichloromethane	9.888	83	492008	144.248	ug/l	99
42) Vinyl Acetate	6.600	43	4294003	791.735	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122024\
 Data File : VN085276.D
 Acq On : 20 Dec 2024 11:42
 Operator : JC\MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC150

Quant Time: Dec 21 01:44:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N122024W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Dec 21 01:29:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.559	43	480322	151.167	ug/l	100
44) Isopropyl Acetate	8.682	43	803196	152.166	ug/l	99
45) 1,4-Dioxane	9.694	88	118082	2876.180	ug/l	98
46) Methyl methacrylate	9.677	41	383334	156.458	ug/l	97
47) n-amyl Acetate	12.494	43	748297	166.415	ug/l	92
48) t-1,3-Dichloropropene	10.835	75	518772	153.981	ug/l	95
49) cis-1,3-Dichloropropene	10.312	75	545222	150.530	ug/l	99
50) 1,1,2-Trichloroethane	11.012	97	297532	139.393	ug/l	97
51) Ethyl methacrylate	10.871	69	537584	166.516	ug/l	93
52) 1,3-Dichloropropane	11.159	76	537291	144.204	ug/l	99
53) Dibromochloromethane	11.359	129	359410	143.703	ug/l	98
54) 1,2-Dibromoethane	11.465	107	305011	147.658	ug/l	99
55) 2-Chloroethyl vinyl ether	10.159	63	1320616	794.138	ug/l	99
56) Bromoform	12.576	173	241680	148.125	ug/l	98
58) 4-Methyl-2-Pentanone	10.441	43	2426748	760.003	ug/l	98
59) 2-Hexanone	11.194	43	1789778	771.576	ug/l	97
61) Tetrachloroethene	11.100	164	243812	138.658	ug/l	94
62) Toluene	10.629	91	1393988	144.057	ug/l	100
64) Chlorobenzene	11.888	112	857578	146.943	ug/l	100
65) 1,1,1,2-Tetrachloroethane	11.959	131	302293	142.827	ug/l	99
66) Ethyl Benzene	11.959	91	1580537	155.229	ug/l	99
67) m/p-Xylenes	12.070	106	1160717	303.865	ug/l	100
68) o-Xylene	12.394	106	556306	152.896	ug/l	98
69) Styrene	12.406	104	965130	156.955	ug/l	99
70) Isopropylbenzene	12.694	105	1453567	159.099	ug/l	99
71) 1,1,2,2-Tetrachloroethane	12.935	83	458872	141.611	ug/l	97
72) 1,2,3-Trichloropropane	12.994	75	372774m	128.375	ug/l	
73) Bromobenzene	12.976	156	334967	148.408	ug/l	97
74) n-propylbenzene	13.035	91	1770285	155.531	ug/l	98
75) 2-Chlorotoluene	13.123	91	1049767	153.101	ug/l	99
76) 1,3,5-Trimethylbenzene	13.170	105	1204028	154.448	ug/l	100
77) t-1,4-Dichloro-2-butene	12.735	75	188674	167.396	ug/l	90
78) 4-Chlorotoluene	13.217	91	1068236	153.228	ug/l	99
79) tert-butylbenzene	13.435	119	1012350	158.790	ug/l	99
80) 1,2,4-Trimethylbenzene	13.482	105	1219847	157.263	ug/l	100
81) sec-Butylbenzene	13.612	105	1463612	159.012	ug/l	98
82) p-Isopropyltoluene	13.729	119	1232499	162.592	ug/l	98
83) 1,3-Dichlorobenzene	13.729	146	628188	149.235	ug/l	98
84) 1,4-Dichlorobenzene	13.812	146	628233	154.265	ug/l	97
85) n-Butylbenzene	14.053	91	1110120	169.627	ug/l	98
86) Hexachloroethane	14.329	117	242863	149.237	ug/l	99
87) 1,2-Dichlorobenzene	14.106	146	598727	149.716	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.717	75	94207	159.257	ug/l	93
89) 1,2,4-Trichlorobenzene	15.841	180	316396	166.998	ug/l	99
90) Hexachlorobutadiene	15.500	225	135495	143.473	ug/l	92
91) Naphthalene	15.641	128	1126942	185.128	ug/l	100
92) 1,2,3-Trichlorobenzene	15.841	180	316396	166.998	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN122024\
Data File : VN085276.D
Acq On : 20 Dec 2024 11:42
Operator : JC\MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC150

Quant Time: Dec 21 01:44:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N122024W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Dec 21 01:29:00 2024
Response via : Initial Calibration

