

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5328

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VN085280.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1220WBSD01	Benzene	20	18.9	ug/L	95	11		65	135	20
	Toluene	20	19.4	ug/L	97	9		70	130	20
	Ethyl Benzene	20	18.7	ug/L	94	11		60	140	20
	m/p-Xylenes	40	38.6	ug/L	97	12		87	111	20
	o-Xylene	20	18.7	ug/L	94	10		87	111	20