

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: P5328 SAS No.: P5328 SDG NO.: P5328
 Lab File ID: VN085273.D Date Analyzed: 12/20/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:29
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	29576	7.81	153203	9.10	143051	11.87
UPPER LIMIT	59152	8.312	306406	9.6	286102	12.365
LOWER LIMIT	14788	7.312	76601.5	8.6	71525.5	11.365
EPA SAMPLE NO.						
001-WILLETTS-PT-BLVD (DEC)	29425	7.81	145458	9.10	136162	11.87
002-35TH-AVE (DEC)	28685	7.81	135357	9.10	132214	11.87
VN1220WBL01	28536	7.81	139194	9.10	126992	11.87
VN1220WBS01	31867	7.81	172468	9.10	157990	11.87
VN1220WBSD01	28666	7.81	153928	9.10	142114	11.87

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.