

Duplicate Sample Summary

Client: Willets Point Asphalt Corporation	SDG No.: P5329
Project: SPDES	Sample ID: P5320-02
Client ID: COMPDUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	368		371		1	0.81		12/19/2024