

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

RunNo.: LB134020

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.502	0.50	100	90-110	12/19/2024
Sample ID: CCV1 Orthophosphate as P	mg/L	0.499	0.5	100	90-110	12/19/2024
Sample ID: CCV2 Orthophosphate as P	mg/L	0.499	0.5	100	90-110	12/19/2024

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

RunNo.: LB134022

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.497	0.50	99	90-110	12/19/2024
Sample ID: CCV1 Phosphorus, Total	mg/L	0.497	0.50	99	90-110	12/19/2024
Sample ID: CCV2 Phosphorus, Total	mg/L	0.494	0.50	99	90-110	12/19/2024

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

RunNo.: LB134087

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	12/26/2024
Sample ID: CCV1 Ammonia as N	mg/L	0.98	1	98	90-110	12/26/2024
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	12/26/2024
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	12/26/2024
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	12/26/2024