



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

RunNo.: LB134020

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Orthophosphate as P	mg/L	< 0.0250	0.0250	U	0.0040	0.05	12/19/2024
Sample ID: CCB1 Orthophosphate as P	mg/L	< 0.0250	0.0250	U	0.0040	0.05	12/19/2024
Sample ID: CCB2 Orthophosphate as P	mg/L	< 0.0250	0.0250	U	0.0040	0.05	12/19/2024

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

RunNo.: LB134022

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Phosphorus, Total	mg/L	0.010	0.0250	J	0.0047	0.05	12/19/2024
Sample ID: CCB1 Phosphorus, Total	mg/L	0.010	0.0250	J	0.0047	0.05	12/19/2024
Sample ID: CCB2 Phosphorus, Total	mg/L	0.010	0.0250	J	0.0047	0.05	12/19/2024

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

RunNo.: LB134087

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/26/2024
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/26/2024
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/26/2024
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/26/2024
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/26/2024