

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

Run No.: LB133994

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133994BS							
BOD5	mg/L	198	198		100	1	84.6-115.4	12/18/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

Run No.: LB134009

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134009BS							
TSS	mg/L	550	594		108	1	90-110	12/19/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

Run No.: LB134020

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134020BS							
Orthophosphate as P	mg/L	0.5	0.50		100	1	90-110	12/19/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

Run No.: LB134030

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134030BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	12/20/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

Run No.: LB134022

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165757BS							
Phosphorus, Total	mg/L	0.50	0.49		99	1	90-110	12/19/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5340

Project: Pre Treatment Plant

Run No.: LB134087

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165803BS							
Ammonia as N	mg/L	1	0.98		98	1	90-110	12/26/2024