

Report of Analysis

Client:	PSEG	Date Collected:	12/18/24
Project:	Kingsland Substation	Date Received:	12/18/24
Client Sample ID:	HT2651	SDG No.:	P5342
Lab Sample ID:	P5342-03	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	88.6
Sample Wt/Vol:	10.05	Units:	g
Soil Aliquot Vol:		Final Vol:	2000 uL
Prep Method :		Test:	EPH_NF

Prep Date :	Date Analyzed :	Prep Batch ID
12/19/24 09:10	12/20/24 11:31	PB165749

Datafile

CAS Number	Parameter	Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS							
Aliphatic C28-C40	Aliphatic C28-C40	537		10	60.6	67.4	mg/kg FE051736.D
Aliphatic C9-C28	Aliphatic C9-C28	11400		100	579	1350	mg/kg FE051737.D
Total AliphaticEPH	Total AliphaticEPH	11900			640	1410	mg/kg
Total EPH	Total EPH	11900			640	1410	mg/kg

* As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

Report of Analysis

Client:	PSEG	Date Collected:	12/18/24
Project:	Kingsland Substation	Date Received:	12/18/24
Client Sample ID:	HT2651	SDG No.:	P5342
Lab Sample ID:	P5342-03	Matrix:	Solid
Analytical Method:	NJEPH	% Solid:	88.6
Sample Wt/Vol:	10.05 Units: g	Final Vol:	2000 uL
Soil Aliquot Vol:	uL	Test:	EPH_NF
Prep Method :			

File ID :	Dilution:	Prep Date :	Date Analyzed :	Prep Batch ID
FE051722.D	1	12/19/24	12/19/24	PB165749

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
Aliphatic C9-C28	Aliphatic C9-C28	10900	E	5.80	13.5	mg/kg
Aliphatic C28-C40	Aliphatic C28-C40	540	E	6.06	6.74	mg/kg
SURROGATES						
3383-33-2	1-chlorooctadecane (SURR)	0.00		40 - 140	0%	SPK: 50
84-15-1	ortho-Terphenyl (SURR)	0.00		40 - 140	0%	SPK: 50

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID:	P5342-03	Acq On:	19 Dec 2024 19:31
Client Sample ID:	HT2651	Operator:	YP\AJ
Data file:	FE051722.D	Misc:	
Instrument:	FID_E	ALS Vial:	21
Dilution Factor:	1	Sample Multiplier:	1.00

Compound	R.T.		Response	Conc	highest_standard	Units
Aliphatic C9-C12	3.381	6.952	12361698	99.086	300	ug/ml
Aliphatic C12-C16	6.953	10.358	878338999	6340	200	ug/ml
Aliphatic C16-C21	10.359	13.700	3764638753	26100	300	ug/ml
Aliphatic C21-C28	13.701	17.331	2189881926	16100	400	ug/ml
Aliphatic C28-C40	17.332	22.209	261221603	2410	600	ug/ml
Aliphatic EPH	3.381	22.209	7106442979	51000		ug/ml
ortho-Terphenyl (SURR)	0.000	0.000	0	0		ug/ml
1-chlorooctadecane (SURR)	0.000	0.000	0	0		ug/ml
Aliphatic C9-C28	3.381	17.331	6845221376	48600	1200	ug/ml