

## Org e Chem

---

**From:** Danganan, Eloise B <Eloise.Danganan@gdit.com>  
**Sent:** Tuesday, December 31, 2024 9:02 AM  
**To:** Mohammad Ahmed; Deepak Parmar; Maheshbhai Dadoda  
**Subject:** REGION 3 | CASE 51914 | LAB ACE | SDGs C0CG7 and C0CF0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL  
**Attachments:** P5309-12(C0CJ1).pdf; P5309-13(C0CJ2).pdf; P5345-02(C0CK0).pdf; P5345-03(C0CK1).pdf; P5345-04(C0CK2).pdf; P5345-05(C0CK3).pdf; P5309-12(C0CJ1).pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning ACE,

Please see the Region's response below regarding the laboratory's reported issue for Case 51914/SDGs C0CG7 and C0CF0. Please let me know if you have any questions.

**ISSUE:** "Lab has analyzed Mutiple samples in a continuous analytical sequence where sample C0CJ1, C0CJ2, C0CK0, C0CK1, C0CK2, C0CK3 had different target analytes detected at very high concentration over the calibration. Due to these samples were analyzed back-to-back in a continuous analytical sequence instrument blanks were not analyzed in between these samples and due to these samples contain high concentration target compounds lab will use this undiluted analysis for our first run and analyze required dilution. Lab will report this undiluted VOA analysis without instrument blank and diluted samples in final electronic deliverables.  
Raw data for samples is attached for your review."

**REGION:** "Inform ACE Region 3 is in agreement with their below written approach; have ACE make note of the issue in their SDG Narrative and proceed with the analysis of the samples."

\*\*\*\*\* Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

ELOISE DANGANAN  
Environmental Analyst  
GDIT Federal Civilian Division  
Under contract to the EPA  
T 703.254.0067  
15036 Conference Center Drive  
Chantilly, VA 20151  
[www.gdit.com](http://www.gdit.com)

**GENERAL DYNAMICS**  
Information Technology

LEAVE ALERT: none

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

---

**From:** Burman, Jarmael <Burman.Jarmael@epa.gov>  
**Sent:** Tuesday, December 31, 2024 8:55 AM  
**To:** Danganan, Eloise B <Eloise.Danganan@gdit.com>  
**Subject:** RE: REGION 3 | CASE 51914 | LAB ACE | SDGs C0CG7 and C0CF0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

**This Message Is From an External Sender**

Please use caution with links, attachments, and any requests for credentials.

Good morning Eloise,

Inform ACE Region 3 is in agreement with their below written approach; have ACE make note of the issue in their SDG Narrative and proceed with the analysis of the samples.

Jarmael Burman  
US EPA Region 3 - CLP RR/RSCC/DDS/QA Chemist/DAS PO/EEOC  
701 Mapes Road  
Fort Meade, Maryland 20755-5350  
(410) 305-2743 (office)  
(410) 305-3095 (fax)

---

**From:** Danganan, Eloise B <[Eloise.Danganan@gdit.com](mailto:Eloise.Danganan@gdit.com)>  
**Sent:** Tuesday, December 31, 2024 8:49 AM  
**To:** Burman, Jarmael <[Burman.Jarmael@epa.gov](mailto:Burman.Jarmael@epa.gov)>  
**Subject:** REGION 3 | CASE 51914 | LAB ACE | SDGs C0CG7 and C0CF0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

**Caution:** This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Jay,

Please advise on the issue below reported by ACE for Case 51914/SDGs C0CG7 and C0CF0.

**ISSUE:** "Lab has analyzed Mutiple samples in a continuous analytical sequence where sample C0CJ1, C0CJ2, C0CK0, C0CK1, C0CK2, C0CK3 had different target analytes detected at very high concentration over the calibration. Due to these samples were analyzed back-to-back in a continuous analytical sequence instrument blanks were not analyzed in between these samples and due to these samples contain high concentration target compounds lab will use this undiluted analysis for our first run and analyze required dilution. Lab will report this undiluted VOA analysis without instrument blank and diluted samples in final electronic deliverables.

Raw data for samples is attached for your review.”

Thank you,

ELOISE DANGANAN  
Environmental Analyst  
GDIT Federal Civilian Division  
Under contract to the EPA  
T 703.254.0067  
15036 Conference Center Drive  
Chantilly, VA 20151  
[www.gdit.com](http://www.gdit.com)

**GENERAL DYNAMICS**  
Information Technology

LEAVE ALERT: none

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

---

**From:** Mohammad Ahmed <[mohammad.ahmed@alliancetg.com](mailto:mohammad.ahmed@alliancetg.com)>  
**Sent:** Monday, December 30, 2024 8:35 PM  
**To:** DeBerry, Eric <[Eric.Deberry@gdit.com](mailto:Eric.Deberry@gdit.com)>; DASsupport <[DASsupport@gdit.com](mailto:DASsupport@gdit.com)>  
**Cc:** Deepak Parmar <[Deepak.Parmar@alliancetg.com](mailto:Deepak.Parmar@alliancetg.com)>; Maheshbhai Dadoda <[Maheshbhai.Dadoda@alliancetg.com](mailto:Maheshbhai.Dadoda@alliancetg.com)>  
**Subject:** REGION 3 | CASE 51914 | LAB ACE | SDG C0CG7, C0CF0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE analysis

**This Message Is From an External Sender**

Please use caution with links, attachments, and any requests for credentials.

Hi,  
Lab has analyzed Mutiple samples in a continuous analytical sequence where sample C0CJ1, C0CJ2, C0CK0, C0CK1, C0CK2, C0CK3 had different target analytes detected at very high concentration over the calibration. Due to these samples were analyzed back-to-back in a continuous analytical sequence instrument blanks were not analyzed in between these samples and due to these samples contain high concentration target compounds lab will use this undiluted analysis for our first run and analyze required dilution. Lab will report this undiluted VOA analysis without instrument blank and diluted samples in final electronic deliverables.  
Raw data for samples is attached for your review.



**Mohammad Ahmed**  
Laboratory Director  
An Alliance Technical Group Company  
Main: 908-789-8900

Direct: 908-728-3151  
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

