

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51918 MA No.: _____ SDG No.: MJNBZ5
Instrument ID: CV1 Start Date: 01/02/2025
Analytical Method: CVAA Run Batch: LB134147
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.033	0	0.2	0.24	22	2.5	2.5	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51918 MA No.: _____ SDG No.: MJNBZ5
Instrument ID: CV1 Start Date: 01/02/2025
Analytical Method: CVAA Run Batch: LB134147
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.8	-3	7.5	7.5	0	10	10	1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51918 MA No.: _____ SDG No.: MJNBZ5
Instrument ID: CV1 Start Date: 01/07/2025
Analytical Method: CVAA Run Batch: LB134177
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.16	-19	2.5	2.7	9

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51918 MA No.: _____ SDG No.: MJNBZ5
Instrument ID: CV1 Start Date: 01/07/2025
Analytical Method: CVAA Run Batch: LB134177
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.9	-3	7.5	7.4	-2	10	10	1