

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51918 MA No. : SDG No.: MJNBZ5
 Initial Calibration Verification Source : MP83845
 Continuing Calibration Verification Source : MP83847
 Run Batch: LB134147 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV040				ID: CCV094				ID: CCV095		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.9	98		5.0	4.8	96		4.7	94	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51918 MA No. : SDG No.: MJNBZ5
 Initial Calibration Verification Source : MP83879
 Continuing Calibration Verification Source : MP83881
 Run Batch: LB134177 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV046				ID: CCV009				ID: CCV010		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	4.1	103		5.0	5.2	104		5.1	101	