

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134219

Review By	Sarabjit Jaswal	Review On	1/9/2025 3:36:57 PM
Supervise By	Mohan Bera	Supervise On	1/9/2025 3:38:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE	TUNE	01/09/25 12:57		Jaswal	OK
2	S00	S0	CAL1	01/09/25 13:29		Jaswal	OK
3	S01	S01	CAL2	01/09/25 13:32		Jaswal	OK
4	S02	S02	CAL3	01/09/25 13:39		Jaswal	OK
5	S03	S03	CAL4	01/09/25 13:42		Jaswal	OK
6	S04	S04	CAL5	01/09/25 13:45		Jaswal	OK
7	S05	S05	CAL6	01/09/25 13:48		Jaswal	OK
8	S06	S06	CAL7	01/09/25 13:51		Jaswal	OK
9	S07	S07	CAL8	01/09/25 13:53		Jaswal	OK
10	S08	S08	CAL9	01/09/25 13:56		Jaswal	OK
11	ICV001	ICV001	ICV	01/09/25 14:04	Instrument paused for calibration review	Jaswal	OK
12	ICB001	ICB001	ICB	01/09/25 14:09		Jaswal	OK
13	ICSA001	ICSA001	ICSA	01/09/25 14:13		Jaswal	OK
14	ICSAB001	ICSAB001	ICSAB	01/09/25 14:16		Jaswal	OK
15	CCV001	CCV001	CCV	01/09/25 14:19		Jaswal	OK
16	CCB001	CCB001	CCB	01/09/25 14:25	Instrument paused for review	Jaswal	OK
17	P5347-05	MJNBZ5	SAM	01/09/25 14:28		Jaswal	OK
18	P5347-06	MJNBZ5D	DUP	01/09/25 14:31		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134219

Review By	Sarabjit Jaswal	Review On	1/9/2025 3:36:57 PM
Supervise By	Mohan Bera	Supervise On	1/9/2025 3:38:47 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

19	P5347-05L	MJNBZ5L	SD	01/09/25 14:34		Jaswal	OK
20	P5347-07	MJNBZ5S	MS	01/09/25 14:37		Jaswal	OK
21	CCV002	CCV002	CCV	01/09/25 14:50		Jaswal	OK
22	CCB002	CCB002	CCB	01/09/25 14:53		Jaswal	OK