
From: Shaeffer, Casey <Casey.Shaeffer@gdit.com>
Sent: Thursday, December 19, 2024 11:56 AM
To: Mohammad Ahmed; Deepak Parmar; Sohil Jodhani
Cc: Johnson, Matthew; Bauer, Heather E; Dunn, Meghan (she/her/hers; Reece, Caitlin; Britz, Helen; Moody, Brett; Gambrah, Derrick; Taylor, Lucinda; Myer, Shari; Patel, Bhavita; Vargas, Magda (she/her/hers
Subject: Region 10 | Case 51918 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC | FINAL
Attachments: SKM_95824121910390.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning,

Please see the below resolution.

Issue: There is no sample designated on the COC for laboratory QC for ICP-MS Metals and Hg analyses in SDG MJNBZ9. The laboratory has selected water sample MJNBZ9 for ICP-MS Metals, water sample MJNC00 for Hg, and soil sample MJNBZ5 for laboratory QC. The laboratory has confirmed that the samples are not a PT, blank, or rinsates.
Resolution: Per SFAM01.1 Exhibit A, Section 5.5.4.1., the laboratory should note the issue in the SDG Narrative and proceed with analysis of the samples.

Please note that the laboratory may contact the appropriate CLP PM should any defects need to be waived for this issue.

Thank you,

Casey Shaeffer

Associate Environmental Analyst
CLP QSS Coordinator – EPA Regions 4 & 10
Under contract to the EPA

T: (571) 454-2416
casey.shaeffer@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: December 24, 2024

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>

Sent: Thursday, December 19, 2024 11:23 AM

To: Shaeffer, Casey <Casey.Shaeffer@gdit.com>

Cc: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>

Subject: Region 10 | Case 51918 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,

Issue 1 : One SDG MJNBZ9 is open without lab QC for ICP-MS and Hg analysis However, a sample was not designated for Laboratory QC. Lab like to use water sample MJNBZ9 for ICP-MS water sample MJNC00 for HG and MJNBZ5 for soil Lab QC. these samples are not blanks, rinsates or PE samples. Case complete.

Please see attachment for your reference.

Thanks & Regards,



Deepak Parmar

QA/QC

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3154

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com

