

Deepak Parmar

From: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>
Sent: Friday, December 20, 2024 4:46 PM
To: Deepak Parmar; Sohil Jodhani; Mohammad Ahmed
Cc: R5RSCC; Bauer, Heather E; Johnson, Matthew; Helen Britz; Moody, Brett; Gambrah, Derrick; vargas.magda@epa.gov; patel.bhavita@epa.gov
Subject: Region 05 | Case 51847 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC | FINAL
Attachments: SKM_95824122012030.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Issue 1: SDG E2927 requires Laboratory QC for water samples, but a sample was not designated on the COC. The laboratory would like to proceed without Laboratory QC to keep holding time.

Resolution 1: Per Region 5, the laboratory will note the issue in the SDG Narrative and proceed without Laboratory QC.

Issue 2: SDG ME2927 requires Laboratory QC for water samples for ICP-MS, Hg, CN analysis, but a sample was not designated on the COC. The laboratory selected sample ME2928 for Laboratory QC and confirmed this sample is not a blank, rinsate or PT sample.

Resolution 2: Per SFAM01.1 Exhibit A, Section 5.5.4.1., the laboratory will note the issue in the SDG Narrative and proceed with the analysis of the samples.

Please note that the laboratory may contact the appropriate CLP PM should any defects need to be waived for this issue.

Best Regards,
Tasmia Zafar
Associate Environmental Analyst
CLP QSS Coordinator – EPA Regions 5 & 6

T: (919) 768-4086
tasmia.zafar@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: 12/23/24 and 12/26/24-12/27/24

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly

prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>

Sent: Friday, December 20, 2024 4:07 PM

To: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>

Subject: RE: Region 05 | Case 51847 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good afternoon Tasmia,

Please have the lab proceed without lab QC.

Thanks,

Darcie D. Tone-Pah-Hote

Regional Sample Control Coordinator
LSASD-SQAB

U.S. EPA Region 5

77 W Jackson Blvd

Chicago, IL 60604

Cubicle #: 06078

Phone #: 312-886-4298

From: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>

Sent: Friday, December 20, 2024 11:53 AM

To: R5RSCC <R5RSCC@epa.gov>

Subject: Region 05 | Case 51847 | Lab ACE | Issue Insufficient/inappropriate designation of laboratory QC

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon,

Please advise on Issue 1 below as Issue 2 has been resolved through the SOW.

Issue 1: SDG E2927 requires Laboratory QC for water samples, but a sample was not designated on the COC. The laboratory would like to proceed without Laboratory QC to keep holding time.

Issue 2: SDG ME2927 requires Laboratory QC for water samples for ICP-MS, Hg, CN analysis, but a sample was not designated on the COC. The laboratory selected sample ME2928 for Laboratory QC and confirmed this sample is not a blank, rinsate or PT sample.

Resolution 2: Per SFAM01.1 Exhibit A, Section 5.5.4.1., the laboratory will note the issue in the SDG Narrative and proceed with the analysis of the samples.

Best Regards,
Tasmia Zafar
Associate Environmental Analyst
CLP QSS Coordinator – EPA Regions 5 & 6

T: (919) 768-4086
tasmia.zafar@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: 12/23/24 and 12/26/24-12/27/24

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>
Sent: Friday, December 20, 2024 12:23 PM
To: Zafar, Tasmia (NE) <Tasmia.Zafar@gdit.com>
Cc: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Subject: Region 05 | Case 51847 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC/QC

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good afternoon,

Issue 1 : One SDG E2927 is open without lab QC for water sample. However, a sample was not designated for Laboratory QC.. The laboratory would like to proceed without Laboratory QC to keep holding time. Also, please confirm that the any SEDD defect associated to Lab QC will be considered Invalid.

Issue 2: Lab has open SDG ME2927 for water sample for ICP-MS,HG,CN without lab QC. However, a sample was not designated for Laboratory QC. Lab like to use samples ME2928 for Lab QC. these samples are not blanks, rinsates or PE samples.

Please see attachment for your reference.

Thanks & Regards,



Deepak Parmar

QA/QC

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3154

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com

