

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME2927
Instrument ID: KONELAB Start Date: 12/27/2024
Analytical Method: Spectrophotometry Run Batch: LB134108
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Cyanide	0	0.29	0	5	5.1	2	10	9.8	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME2927
Instrument ID: KONELAB Start Date: 12/27/2024
Analytical Method: Spectrophotometry Run Batch: LB134108
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Cyanide	100	100	0	250	250	0	500	500	0