

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME2927
 Instrument ID: P7 Start Date : 12/27/2024
 Run Batch: LB134112 End Date : 12/27/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
S0	1436	100		100		100		100		100	
S01	1446	101		101		102		102		101	
S02	1449	103		104		104		103		102	
S03	1452	104		102		102		102		103	
S04	1455	102		96		99		97		102	
S05	1458	103		93		97		94		102	
S06	1501	99		88		93		90		99	
S07	1504	98		86		92		86		99	
S08	1507	95		87		90		81		96	
ICV002	1530	103		95		97		96		100	
ICB002	1538	104		96		98		97		99	
ICSA002	1542	98		88		94		87		99	
ICSAB002	1545	98		93		98		92		104	
CCV011	1548	96		90		95		87		101	
CCB011	1551	99		88		93		91		97	
ZZZZZZ	1557										
ZZZZZZ	1601										
ZZZZZZ	1604										
ZZZZZZ	1607										
ZZZZZZ	1610										
ZZZZZZ	1613										
ZZZZZZ	1616										
ZZZZZZ	1619										
CCV012	1623	95		92		96		88		102	
CCB012	1625	104		95		100		99		106	
ZZZZZZ	1629										
ZZZZZZ	1632										
ZZZZZZ	1635										
ZZZZZZ	1641										
ZZZZZZ	1644										
ZZZZZZ	1647										
ZZZZZZ	1650										
ZZZZZZ	1653										
ZZZZZZ	1657										
ZZZZZZ	1700										
ZZZZZZ	1703										
ZZZZZZ	1706										
ZZZZZZ	1709										

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
ZZZZZZ	1712										
ZZZZZZ	1716										
ZZZZZZ	1720										
ZZZZZZ	1723										
ZZZZZZ	1726										
ZZZZZZ	1729										
ZZZZZZ	1733										
ZZZZZZ	1736										
ZZZZZZ	1739										
ZZZZZZ	1742										
ZZZZZZ	1745										
ZZZZZZ	1749										
PBW859	1752	101		100		102		102		103	
LCS859	1755	103		102		105		105		106	
CCV013	1801	97		89		93		86		101	
CCB013	1811	104		95		100		99		105	
ZZZZZZ	1814										
ZZZZZZ	1818										
ZZZZZZ	1821										
ZZZZZZ	1824										
ZZZZZZ	1828										
ZZZZZZ	1831										
ZZZZZZ	1834										
ZZZZZZ	1837										
ZZZZZZ	1840										
ZZZZZZ	1843										
ZZZZZZ	1846										
ZZZZZZ	1849										
ZZZZZZ	1852										
ZZZZZZ	1856										
ZZZZZZ	1859										
ZZZZZZ	1902										
ZZZZZZ	1905										
ZZZZZZ	1908										
ZZZZZZ	1911										
ZZZZZZ	1915										
ZZZZZZ	1918										
CCV014	1921	96		91		95		88		102	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME2927
 Instrument ID: P7 Start Date : 12/27/2024
 Run Batch: LB134112 End Date : 12/27/2024

EPA Sample No.	Time	Internal Standard %RI For: Non-Collision Cell									
		Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
		6Li		45Sc		89Y		103Rh		159Tb	
CCB014	1924	99		90		97		97		104	
ZZZZZZ	1927										
ZZZZZZ	1931										
ZZZZZZ	1934										
ME2928	1937	99		93		99		99		104	
ME2928D	1940	99		92		99		97		104	
ME2928L	1944	99		93		99		99		103	
ME2928S	1947	101		94		100		99		105	
ZZZZZZ	1950										
ZZZZZZ	1954										
ZZZZZZ	1957										
ME2927	2000	96		93		101		100		106	
ZZZZZZ	2004										
ZZZZZZ	2007										
ZZZZZZ	2010										
ZZZZZZ	2013										
ZZZZZZ	2017										
ZZZZZZ	2020										
ZZZZZZ	2023										
ZZZZZZ	2026										
ZZZZZZ	2029										
ZZZZZZ	2032										
ZZZZZZ	2036										
ZZZZZZ	2039										
ZZZZZZ	2042										
ZZZZZZ	2045										
ZZZZZZ	2049										
ZZZZZZ	2052										
CCV015	2055	93		88		94		87		102	
CCB015	2106	99		93		100		99		105	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME2927
 Instrument ID: P7 Start Date : 12/27/2024
 Run Batch: LB134112 End Date : 12/27/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
S0	1436	100		100		100		100		100	
S01	1446	103		103		103		102		101	
S02	1449	102		102		103		102		101	
S03	1452	98		99		99		100		100	
S04	1455	92		94		95		99		99	
S05	1458	86		90		90		98		98	
S06	1501	82		87		87		97		97	
S07	1504	82		87		85		97		97	
S08	1507	87		88		80		93		94	
ICV002	1530	89		92		94		97		97	
ICB002	1538	96		96		98		99		98	
ICSA002	1542	87		92		89		100		100	
ICSAB002	1545	90		93		91		101		101	
CCV011	1548	88		91		86		98		98	
CCB011	1551	93		96		99		103		102	
ZZZZZZ	1557										
ZZZZZZ	1601										
ZZZZZZ	1604										
ZZZZZZ	1607										
ZZZZZZ	1610										
ZZZZZZ	1613										
ZZZZZZ	1616										
ZZZZZZ	1619										
CCV012	1623	88		91		86		98		98	
CCB012	1625	92		96		98		102		102	
ZZZZZZ	1629										
ZZZZZZ	1632										
ZZZZZZ	1635										
ZZZZZZ	1641										
ZZZZZZ	1644										
ZZZZZZ	1647										
ZZZZZZ	1650										
ZZZZZZ	1653										
ZZZZZZ	1657										
ZZZZZZ	1700										
ZZZZZZ	1703										
ZZZZZZ	1706										
ZZZZZZ	1709										

ICP-MS	INTERNAL	STANDARD	RELATIVE	INTENSITY	SUMMARY
--------	----------	----------	----------	-----------	---------

Contract: 68HERH20D0011

MA No.: SDG No.: ME2927

Start Date : 12/27/2024

End Date : 12/27/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
ZZZZZZ	1712										
ZZZZZZ	1716										
ZZZZZZ	1720										
ZZZZZZ	1723										
ZZZZZZ	1726										
ZZZZZZ	1729										
ZZZZZZ	1733										
ZZZZZZ	1736										
ZZZZZZ	1739										
ZZZZZZ	1742										
ZZZZZZ	1745										
ZZZZZZ	1749										
PBW859	1752	99		100		102		102		102	
LCS859	1755	100		101		102		102		102	
CCV013	1801	85		89		85		98		99	
CCB013	1811	92		96		99		102		102	
ZZZZZZ	1814										
ZZZZZZ	1818										
ZZZZZZ	1821										
ZZZZZZ	1824										
ZZZZZZ	1828										
ZZZZZZ	1831										
ZZZZZZ	1834										
ZZZZZZ	1837										
ZZZZZZ	1840										
ZZZZZZ	1843										
ZZZZZZ	1846										
ZZZZZZ	1849										
ZZZZZZ	1852										
ZZZZZZ	1856										
ZZZZZZ	1859										
ZZZZZZ	1902										
ZZZZZZ	1905										
ZZZZZZ	1908										
ZZZZZZ	1911										
ZZZZZZ	1915										
ZZZZZZ	1918										
CCV014	1921	84		88		84		97		97	

FORM 14-IN

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME2927
 Instrument ID: P7 Start Date : 12/27/2024
 Run Batch: LB134112 End Date : 12/27/2024

EPA Sample No.	Time	Internal Standard %RI For: Collision Cell									
		Element		Element		Element		Element		Element	
		45Sc	Q	89Y	Q	103Rh	Q	159Tb	Q	165Ho	Q
CCB014	1924	87		93		96		101		100	
ZZZZZZ	1927										
ZZZZZZ	1931										
ZZZZZZ	1934										
ME2928	1937	89		94		97		101		101	
ME2928D	1940	89		94		97		100		101	
ME2928L	1944	89		95		97		100		101	
ME2928S	1947	91		95		98		101		101	
ZZZZZZ	1950										
ZZZZZZ	1954										
ZZZZZZ	1957										
ME2927	2000	88		94		98		102		103	
ZZZZZZ	2004										
ZZZZZZ	2007										
ZZZZZZ	2010										
ZZZZZZ	2013										
ZZZZZZ	2017										
ZZZZZZ	2020										
ZZZZZZ	2023										
ZZZZZZ	2026										
ZZZZZZ	2029										
ZZZZZZ	2032										
ZZZZZZ	2036										
ZZZZZZ	2039										
ZZZZZZ	2042										
ZZZZZZ	2045										
ZZZZZZ	2049										
ZZZZZZ	2052										
CCV015	2055	82		87		85		97		98	
CCB015	2106	89		94		98		101		101	