284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM **Case No.:** P5361 **SAS No.:** P5361 **SDG NO.:** P5361

Continuing Calib Date: 12/20/2024 **Initial Calibration Date(s):** 12/06/2024 12/06/2024

Continuing Calib Time: 09:19 **Initial Calibration Time(s):** 14:19 22:34

GC Column: ZB-MR1 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1242-1	(1)	4.80	4.81	4.71	4.91	0.01
Aroclor-1242-2	(2)	4.82	4.83	4.73	4.93	0.01
Aroclor-1242-3	(3)	4.88	4.88	4.78	4.98	0.00
Aroclor-1242-4	(4)	5.00	5.01	4.91	5.11	0.01
Aroclor-1242-5	(5)	5.65	5.66	5.56	5.76	0.01
Tetrachloro-m-xylene		3.71	3.71	3.61	3.81	0.01
Decachlorobiphenyl		8.77	8.79	8.69	8.89	0.02

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM **Case No.:** P5361 **SAS No.:** P5361 **SDG NO.:** P5361

Continuing Calib Date: 12/20/2024 **Initial Calibration Date(s):** 12/06/2024 12/06/2024

Continuing Calib Time: 09:19 **Initial Calibration Time(s):** 14:19 22:34

GC Column: ZB-MR2 **ID:** 0.32 **(mm)**

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM	RT WINDOW TO	DIFF RT
Aroclor-1242-1	(1)	4.79	4.80	4.70	4.90	0.01
Aroclor-1242-2	(2)	4.81	4.82	4.72	4.92	0.01
Aroclor-1242-3	(3)	4.98	4.99	4.89	5.09	0.01
Aroclor-1242-4	(4)	5.07	5.08	4.98	5.18	0.01
Aroclor-1242-5	(5)	5.59	5.60	5.50	5.70	0.01
Tetrachloro-m-xylene		3.70	3.71	3.61	3.81	0.01
Decachlorobiphenyl		8.72	8.74	8.64	8.84	0.02