

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5361 SAS No.: P5361 SDG NO.: P5361

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108695.D Time Analyzed: 09:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.802	4.709	4.909	540.200	500.000	8.0
Aroclor-1016-2	4.822	4.729	4.929	555.450	500.000	11.1
Aroclor-1016-3	4.878	4.785	4.985	556.300	500.000	11.3
Aroclor-1016-4	4.999	4.907	5.107	553.360	500.000	10.7
Aroclor-1016-5	5.257	5.165	5.365	537.750	500.000	7.6
Aroclor-1260-1	6.300	6.210	6.410	526.700	500.000	5.3
Aroclor-1260-2	6.489	6.398	6.598	514.670	500.000	2.9
Aroclor-1260-3	6.858	6.769	6.969	492.010	500.000	-1.6
Aroclor-1260-4	7.119	7.029	7.229	506.910	500.000	1.4
Aroclor-1260-5	7.360	7.270	7.470	520.510	500.000	4.1
Decachlorobiphenyl	8.775	8.691	8.891	48.660	50.000	-2.7
Tetrachloro-m-xylene	3.705	3.610	3.810	58.880	50.000	17.8

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5361 SAS No.: P5361 SDG NO.: P5361

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL01 Date Analyzed: 12/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108695.D Time Analyzed: 09:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.788	4.698	4.898	572.050	500.000	14.4
Aroclor-1016-2	4.809	4.718	4.918	577.950	500.000	15.6
Aroclor-1016-3	4.984	4.894	5.094	566.540	500.000	13.3
Aroclor-1016-4	5.026	4.935	5.135	551.940	500.000	10.4
Aroclor-1016-5	5.240	5.150	5.350	576.040	500.000	15.2
Aroclor-1260-1	6.274	6.186	6.386	563.160	500.000	12.6
Aroclor-1260-2	6.461	6.373	6.573	567.390	500.000	13.5
Aroclor-1260-3	6.615	6.527	6.727	557.150	500.000	11.4
Aroclor-1260-4	7.087	7.000	7.200	566.280	500.000	13.3
Aroclor-1260-5	7.327	7.239	7.439	583.460	500.000	16.7
Decachlorobiphenyl	8.723	8.641	8.841	54.460	50.000	8.9
Tetrachloro-m-xylene	3.702	3.608	3.808	58.010	50.000	16.0