

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5361 SAS No.: P5361 SDG NO.: P5361

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108708.D Time Analyzed: 14:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.802	4.709	4.909	513.110	500.000	2.6
Aroclor-1016-2	4.821	4.729	4.929	445.990	500.000	-10.8
Aroclor-1016-3	4.879	4.785	4.985	489.170	500.000	-2.2
Aroclor-1016-4	5.000	4.907	5.107	482.600	500.000	-3.5
Aroclor-1016-5	5.257	5.165	5.365	464.390	500.000	-7.1
Aroclor-1260-1	6.300	6.210	6.410	466.240	500.000	-6.8
Aroclor-1260-2	6.489	6.398	6.598	463.300	500.000	-7.3
Aroclor-1260-3	6.858	6.769	6.969	453.360	500.000	-9.3
Aroclor-1260-4	7.119	7.029	7.229	447.240	500.000	-10.6
Aroclor-1260-5	7.360	7.270	7.470	445.550	500.000	-10.9
Decachlorobiphenyl	8.775	8.691	8.891	43.650	50.000	-12.7
Tetrachloro-m-xylene	3.706	3.610	3.810	47.630	50.000	-4.7

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5361 SAS No.: P5361 SDG NO.: P5361

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL02 Date Analyzed: 12/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108708.D Time Analyzed: 14:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.790	4.698	4.898	515.410	500.000	3.1
Aroclor-1016-2	4.810	4.718	4.918	513.630	500.000	2.7
Aroclor-1016-3	4.986	4.894	5.094	508.290	500.000	1.7
Aroclor-1016-4	5.027	4.935	5.135	499.210	500.000	-0.2
Aroclor-1016-5	5.241	5.150	5.350	522.770	500.000	4.6
Aroclor-1260-1	6.276	6.186	6.386	512.750	500.000	2.6
Aroclor-1260-2	6.463	6.373	6.573	511.890	500.000	2.4
Aroclor-1260-3	6.616	6.527	6.727	511.380	500.000	2.3
Aroclor-1260-4	7.088	7.000	7.200	531.830	500.000	6.4
Aroclor-1260-5	7.329	7.239	7.439	533.750	500.000	6.8
Decachlorobiphenyl	8.726	8.641	8.841	49.680	50.000	-0.6
Tetrachloro-m-xylene	3.704	3.608	3.808	51.410	50.000	2.8