

### CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5361 SAS No.: P5361 SDG NO.: P5361

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108723.D Time Analyzed: 20:11

| COMPOUND             | RT    | RT WINDOW<br>FROM TO |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|----------------------|-------|--------------------|-------------------|-------|
| Aroclor-1016-1       | 4.803 | 4.709                | 4.909 | 474.430            | 500.000           | -5.1  |
| Aroclor-1016-2       | 4.822 | 4.729                | 4.929 | 487.550            | 500.000           | -2.5  |
| Aroclor-1016-3       | 4.878 | 4.785                | 4.985 | 489.360            | 500.000           | -2.1  |
| Aroclor-1016-4       | 4.999 | 4.907                | 5.107 | 483.220            | 500.000           | -3.4  |
| Aroclor-1016-5       | 5.257 | 5.165                | 5.365 | 478.880            | 500.000           | -4.2  |
| Aroclor-1260-1       | 6.300 | 6.210                | 6.410 | 485.350            | 500.000           | -2.9  |
| Aroclor-1260-2       | 6.489 | 6.398                | 6.598 | 485.740            | 500.000           | -2.9  |
| Aroclor-1260-3       | 6.858 | 6.769                | 6.969 | 466.730            | 500.000           | -6.7  |
| Aroclor-1260-4       | 7.119 | 7.029                | 7.229 | 458.660            | 500.000           | -8.3  |
| Aroclor-1260-5       | 7.360 | 7.270                | 7.470 | 459.400            | 500.000           | -8.1  |
| Decachlorobiphenyl   | 8.774 | 8.691                | 8.891 | 44.350             | 50.000            | -11.3 |
| Tetrachloro-m-xylene | 3.706 | 3.610                | 3.810 | 50.250             | 50.000            | 0.5   |

### CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5361 SAS No.: P5361 SDG NO.: P5361

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108723.D Time Analyzed: 20:11

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.790 | 4.698     | 4.898 | 528.870            | 500.000           | 5.8  |
| Aroclor-1016-2       | 4.810 | 4.718     | 4.918 | 532.860            | 500.000           | 6.6  |
| Aroclor-1016-3       | 4.986 | 4.894     | 5.094 | 522.170            | 500.000           | 4.4  |
| Aroclor-1016-4       | 5.027 | 4.935     | 5.135 | 504.670            | 500.000           | 0.9  |
| Aroclor-1016-5       | 5.241 | 5.150     | 5.350 | 550.480            | 500.000           | 10.1 |
| Aroclor-1260-1       | 6.275 | 6.186     | 6.386 | 530.890            | 500.000           | 6.2  |
| Aroclor-1260-2       | 6.462 | 6.373     | 6.573 | 532.130            | 500.000           | 6.4  |
| Aroclor-1260-3       | 6.616 | 6.527     | 6.727 | 527.280            | 500.000           | 5.5  |
| Aroclor-1260-4       | 7.088 | 7.000     | 7.200 | 546.940            | 500.000           | 9.4  |
| Aroclor-1260-5       | 7.328 | 7.239     | 7.439 | 547.510            | 500.000           | 9.5  |
| Decachlorobiphenyl   | 8.724 | 8.641     | 8.841 | 50.810             | 50.000            | 1.6  |
| Tetrachloro-m-xylene | 3.703 | 3.608     | 3.808 | 52.800             | 50.000            | 5.6  |