

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120624\
 Data File : PO108381.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2024 20:08
 Operator : YP/AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 04:55:12 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 04:53:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.709	3.707	474.2E6	268.4E6	50.000	50.000
2)	SA Decachlor...	8.786	8.738	392.3E6	212.9E6	50.000	50.000
Target Compounds							
26)	L6 AR-1254-1	5.619	5.601	256.0E6	134.6E6	500.000	500.000
27)	L6 AR-1254-2	5.768	5.749	226.4E6	119.4E6	500.000	500.000
28)	L6 AR-1254-3	6.174	6.153	363.2E6	190.2E6	500.000	500.000
29)	L6 AR-1254-4	6.404	6.381	222.1E6	110.3E6	500.000	500.000
30)	L6 AR-1254-5	6.825	6.799	318.0E6	163.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO120624\
Data File : PO108381.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2024 20:08
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 04:55:12 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO120624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 04:53:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.709 min
Delta R.T.: 0.000 min
Response: 474182655
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.707 min
Delta R.T.: 0.000 min
Response: 268363315
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.786 min
Delta R.T.: 0.000 min
Response: 392292110
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.738 min
Delta R.T.: 0.000 min
Response: 212946579
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 5.619 min
Delta R.T.: 0.000 min
Response: 256015757
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.601 min
Delta R.T.: 0.000 min
Response: 134614973
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 5.768 min
Delta R.T.: 0.000 min
Response: 226446265
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 119385210
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.174 min
Delta R.T.: 0.000 min
Response: 363236901
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.153 min
Delta R.T.: 0.000 min
Response: 190220530
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 6.404 min
Delta R.T.: 0.000 min
Response: 222134078
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.381 min
Delta R.T.: 0.000 min
Response: 110254180
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 6.825 min
Delta R.T.: 0.000 min
Response: 318046447
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.799 min
Delta R.T.: 0.000 min
Response: 163507863
Conc: 500.00 ng/ml