

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P5361

Client: PARSONS Engineering of New York, In

Analytical Method: 8082A

DataFile : PO108714.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: P5355-01MS	RBR251688MS											
	AR1016	203.9	0	232	ug/kg	114				55	146	
	AR1260	203.9	0	232	ug/kg	114				45	144	