



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5361

Client: PARSONS Engineering of New York, In

Analytical Method: 8082A Datafile : PO108707.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165777BS	AR1016	166.5	161	ug/kg	97				71	120	
	AR1260	166.5	164	ug/kg	98				65	130	