



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**PARSONS Main of New York, Inc.**

**Project Name: Con Ed Non-MGP - East River SI 453648**

**Project # N/A**

**Chemtech Project # P5362**

**Test Name: TCLP BNA**

### **A. Number of Samples and Date of Receipt:**

2 Solid samples were received on 12/19/2024.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Corrosivity, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TPH GC and VOC-TCLVOA-10. This data package contains results for TCLP BNA.

### **C. Analytical Techniques:**

The samples were analyzed on instrument BNA\_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df The analysis of TCLP BNA was based on method 8270E and extraction was done based on method 3510 and TCLP extraction method was 1311.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for WC-SOIL-20241219MS [Terphenyl-d14 - 126%], due to matrix interference no corrective action is required.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {P5362-02MS} with File ID: BF141050.D recoveries met the requirements for all compounds except for 2,4,5-Trichlorophenol[114%] and 2,4,6-Trichlorophenol[114%] due to matrix interference no corrective action is required.

The MSD recoveries met the acceptable requirements .

The RPD met criteria .

The Blank Spike for {PB165894BS} with File ID: BF141021.D met requirements for all samples except for 2,4,5-Trichlorophenol[107%], Hexachlorobenzene[109%], The associate samples have no positive hit for these compounds therefore no corrective action was taken.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (8270-BF122624.M) for 2,4-Dinitrotoluene, this compound is passing on Linear Regression.

The Continuous Calibration File ID BF141047.D met the requirements except for 2,4-Dinitrotoluene, The associate samples have no positive hit for these compounds therefore no corrective action was taken.

The Tuning criteria met requirements.

**E. Additional Comments:**

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_