



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: PARSONS Main of New York, Inc.

SDG No.: P5362

Project: Con Ed Non-MGP - East River SI 453648

RunNo.: LB134062

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	12/23/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/23/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	12/23/2024

Initial and Continuing Calibration Verification

Client: PARSONS Main of New York, Inc.

SDG No.: P5362

Project: Con Ed Non-MGP - East River SI 453648

RunNo.: LB134092

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	12/26/2024
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	12/26/2024
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	12/26/2024

Initial and Continuing Calibration Verification

Client: PARSONS Main of New York, Inc.

SDG No.: P5362

Project: Con Ed Non-MGP - East River SI 453648

RunNo.: LB134110

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.00	7	100	90-110	12/27/2024
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	12/27/2024
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	12/27/2024