



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: PARSONS Main of New York, Inc.

SDG No.: P5362

Contract: PARS02

Lab Code: CHEM

Case No.: P5362

SAS No.: P5362

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	97	40 - 160	CV	12/20/2024	14:22	LB134050
CRI01	Aluminum	93.7	100	94	40 - 160	P	12/30/2024	14:32	LB134129
	Antimony	46.6	50.0	93	40 - 160	P	12/30/2024	14:32	LB134129
	Arsenic	21.1	20.0	105	40 - 160	P	12/30/2024	14:32	LB134129
	Barium	91.3	100	91	40 - 160	P	12/30/2024	14:32	LB134129
	Beryllium	5.68	6.0	95	40 - 160	P	12/30/2024	14:32	LB134129
	Cadmium	6.07	6.0	101	40 - 160	P	12/30/2024	14:32	LB134129
	Calcium	1900	2000	95	40 - 160	P	12/30/2024	14:32	LB134129
	Chromium	9.09	10.0	91	40 - 160	P	12/30/2024	14:32	LB134129
	Cobalt	29.0	30.0	96	40 - 160	P	12/30/2024	14:32	LB134129
	Copper	20.6	20.0	103	40 - 160	P	12/30/2024	14:32	LB134129
	Iron	95.3	100	95	40 - 160	P	12/30/2024	14:32	LB134129
	Lead	11.3	12.0	94	40 - 160	P	12/30/2024	14:32	LB134129
	Magnesium	2030	2000	101	40 - 160	P	12/30/2024	14:32	LB134129
	Manganese	19.2	20.0	96	40 - 160	P	12/30/2024	14:32	LB134129
	Nickel	38.7	40.0	97	40 - 160	P	12/30/2024	14:32	LB134129
	Potassium	1950	2000	98	40 - 160	P	12/30/2024	14:32	LB134129
	Selenium	17.9	20.0	89	40 - 160	P	12/30/2024	14:32	LB134129
	Silver	10.4	10.0	104	40 - 160	P	12/30/2024	14:32	LB134129
	Sodium	1930	2000	96	40 - 160	P	12/30/2024	14:32	LB134129
	Thallium	37.1	40.0	93	40 - 160	P	12/30/2024	14:32	LB134129
	Vanadium	37.7	40.0	94	40 - 160	P	12/30/2024	14:32	LB134129
	Zinc	41.1	40.0	103	40 - 160	P	12/30/2024	14:32	LB134129