


LB134125

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.068	0.068	311	0.000	311	0				1
0.2	0.200	0.194	-0.006	862	0.0 %	862					1
2.5	2.500	2.455	-0.045	10682	0.0 %	10682					1
5.0	5.000	4.828	-0.172	20984	0.0 %	20984					1
7.5	7.500	7.783	0.283	33822	0.0 %	33822					1
10.0	10.000	9.872	-0.128	42891	0.0 %	42891					1

LB134125 INSTRUMENT ID : CV1

Reviewed By:mohan
On:12/30/2024
10:39:44 PM
Inst Id :CV1
LB :LB134125

Method: 7470A

Operator: Admin

Date of Analysis: 30 Dec 2024 13:49:42

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	311	-	0.000	7470A	PPB	30 Dec 2024 13:56:35	S	Std
0.2 - 1	50.2	862	-	0.200	7470A	PPB	30 Dec 2024 13:58:52	S	Std
2.5 - 1	52.5	10682	-	2.500	7470A	PPB	30 Dec 2024 14:01:09	S	Std
5.0 - 1	55	20984	-	5.000	7470A	PPB	30 Dec 2024 14:06:03	S	Std
7.5 - 1	57.5	33822	-	7.500	7470A	PPB	30 Dec 2024 14:16:30	S	Std
10.0 - 1	59	42891	-	10.000	7470A	PPB	30 Dec 2024 14:25:31	S	Std
ICV38 - 1	ICV38	16249	3.7373	-	7470A	PPB	30 Dec 2024 14:28:17	U	SMPL
ICB38 - 1	ICB38	-213	-0.0531	-	7470A	PPB	30 Dec 2024 14:30:33	U	SMPL
CCV88 - 1	CCV88	21163	4.8687	-	7470A	PPB	30 Dec 2024 14:32:50	U	SMPL
CCB88 - 1	CCB88	-397	-0.0955	-	7470A	PPB	30 Dec 2024 14:35:05	U	SMPL
CRA - 1	CRA	857	0.1933	-	7470A	PPB	30 Dec 2024 14:37:22	U	SMPL
HighStd - 1	HighStd	43142	9.9294	-	7470A	PPB	30 Dec 2024 14:39:37	U	SMPL
ChkStd - 1	ChkStd	30668	7.0572	-	7470A	PPB	30 Dec 2024 14:41:52	U	SMPL
PB165912BL - 1	PBW	-1	-0.0043	-	7470A	PPB	30 Dec 2024 14:44:11	U	SMPL
P5362-02 - 1	WC-SOIL-20241219	-148	-0.0381	-	7470A	PPB	30 Dec 2024 14:48:45	U	SMPL
P5380-02 - 1	TAPIAL3-IDW-SOIL-122024-T1	85	0.0155	-	7470A	PPB	30 Dec 2024 14:51:04	U	SMPL
P5386-02 - 1	MOO-24-00398	80	0.0144	-	7470A	PPB	30 Dec 2024 14:53:21	U	SMPL
P5386-04 - 1	MOO-24-00395-96	485	0.1076	-	7470A	PPB	30 Dec 2024 15:00:46	U	SMPL
P5386-04DUP - 1	MOO-24-00395-96DUP	459	0.1016	-	7470A	PPB	30 Dec 2024 15:03:00	U	SMPL
CCV89 - 1	CCV89	21149	4.8655	-	7470A	PPB	30 Dec 2024 15:05:15	U	SMPL
CCB89 - 1	CCB89	-185	-0.0466	-	7470A	PPB	30 Dec 2024 15:07:30	U	SMPL
P5386-04MS - 1	MOO-24-00395-96MS	17663	4.0628	-	7470A	PPB	30 Dec 2024 15:09:47	U	SMPL
P5386-04MSD - 1	MOO-24-00395-96MSD	18459	4.2461	-	7470A	PPB	30 Dec 2024 15:12:02	U	SMPL
PB165858TB - 1	PB165858TB	-66	-0.0192	-	7470A	PPB	30 Dec 2024 15:14:19	U	SMPL
P5386-04LX5 - 1		193	0.0404	-	7470A	PPB	30 Dec 2024 15:16:36	U	SMPL
P5386-04A - 1		19290	4.4375	-	7470A	PPB	30 Dec 2024 15:18:52	U	SMPL
PB165912BS - 1	LCSW	16357	3.7621	-	7470A	PPB	30 Dec 2024 15:22:26	U	SMPL
CCV90 - 1	CCV90	21325	4.9060	-	7470A	PPB	30 Dec 2024 15:24:59	U	SMPL
CCB90 - 1	CCB90	181	0.0376	-	7470A	PPB	30 Dec 2024 15:27:14	U	SMPL