

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: PARSONS Main of New York, Inc. **SDG No.:** P5362
Contract: PARS02 **Lab Code:** CHEM **Case No.:** P5362 **SAS No.:** P5362

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165887BS							
Arsenic	ug/L	4000	3890		97	80 - 120	P
Barium	ug/L	1000	922		92	80 - 120	P
Cadmium	ug/L	1000	976		98	80 - 120	P
Chromium	ug/L	2000	2070		104	80 - 120	P
Lead	ug/L	5000	4870		97	80 - 120	P
Selenium	ug/L	10000	9410		94	80 - 120	P
Silver	ug/L	380	371		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: PARSONS Main of New York, Inc.

SDG No.: P5362

Contract: PARS02

Lab Code: CHEM

Case No.: P5362

SAS No.: P5362

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165912BS Mercury	ug/L	4.0	3.76		94	82 - 119	CV