

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108723.D Time Analyzed: 20:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.803	4.709	4.909	474.430	500.000	-5.1
Aroclor-1016-2	4.822	4.729	4.929	487.550	500.000	-2.5
Aroclor-1016-3	4.878	4.785	4.985	489.360	500.000	-2.1
Aroclor-1016-4	4.999	4.907	5.107	483.220	500.000	-3.4
Aroclor-1016-5	5.257	5.165	5.365	478.880	500.000	-4.2
Aroclor-1260-1	6.300	6.210	6.410	485.350	500.000	-2.9
Aroclor-1260-2	6.489	6.398	6.598	485.740	500.000	-2.9
Aroclor-1260-3	6.858	6.769	6.969	466.730	500.000	-6.7
Aroclor-1260-4	7.119	7.029	7.229	458.660	500.000	-8.3
Aroclor-1260-5	7.360	7.270	7.470	459.400	500.000	-8.1
Decachlorobiphenyl	8.774	8.691	8.891	44.350	50.000	-11.3
Tetrachloro-m-xylene	3.706	3.610	3.810	50.250	50.000	0.5

CALIBRATION VERIFICATION SUMMARY

Contract: PARS02

Lab Code: CHEM Case No.: P5362 SAS No.: P5362 SDG NO.: P5362

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/06/2024 12/06/2024

Client Sample No.: CCAL03 Date Analyzed: 12/20/2024

Lab Sample No.: AR1660CCC500 Data File : PO108723.D Time Analyzed: 20:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.790	4.698	4.898	528.870	500.000	5.8
Aroclor-1016-2	4.810	4.718	4.918	532.860	500.000	6.6
Aroclor-1016-3	4.986	4.894	5.094	522.170	500.000	4.4
Aroclor-1016-4	5.027	4.935	5.135	504.670	500.000	0.9
Aroclor-1016-5	5.241	5.150	5.350	550.480	500.000	10.1
Aroclor-1260-1	6.275	6.186	6.386	530.890	500.000	6.2
Aroclor-1260-2	6.462	6.373	6.573	532.130	500.000	6.4
Aroclor-1260-3	6.616	6.527	6.727	527.280	500.000	5.5
Aroclor-1260-4	7.088	7.000	7.200	546.940	500.000	9.4
Aroclor-1260-5	7.328	7.239	7.439	547.510	500.000	9.5
Decachlorobiphenyl	8.724	8.641	8.841	50.810	50.000	1.6
Tetrachloro-m-xylene	3.703	3.608	3.808	52.800	50.000	5.6